



**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [eptenders.kmscl@kerala.gov.in](mailto:eptenders.kmscl@kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

**EP Division**

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

**No: KMSCL/EP/11987/2023**

**Dated: 03.11.2023**

**NOTICE**

**Sub:-** Tender No.T498/ 2023 - Cancellation Notice – Advanced Fluorescent Microscope - reg.

**Ref:-** 1. Tender Inviting Notice published dated 05.07.2023

2. Technical bid opening held on 07.08.2023 & 08.08.2023 @ 11.00 A M @ 11.00 A M

3. Eligible list published on 20.09.2023

4. Reports of Demonstration/ document evaluation of Advanced Fluorescent Microscope received on 31.10.2023

I. The reports of technical document evaluation/ demonstration of Advanced Fluorescent Microscope as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models
1.	Advanced Fluorescent Microscope	M/s Bions Medical Systems Pvt Ltd	Leica DM 2500/ Leica Microsystems

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below:

1. M/s Scientific Medical Equipments (Kerala) Pvt.

Rejected model: AXIOSCOPE 5/ ZEISS CARL ZEISS INDIA

Reasons for rejection: Sample not produced

2. M/s Universal Agencies

Rejected model: BS-2082F/ Beijing BestScope

Reason for rejection: Sample not produced

3. M/s Radical Scientific Equipments Pvt. Ltd

Rejected model: RXLr-5 (Code: NX/ Radical Scientific

Reason for rejection:

Clause No	Technical specification	Compliance of the model
5	Should have 3- step focusing, coarse, medium and fine (1 and 4 microns)	Not meeting the specification
6	Five focusing functions (Coarse, medium, fine, focus height stops, adjustable torque)	Not meeting the specification
9	Condenser Achr.apl.A 0.9 (P) with switchable condenser head, with color coding for fast and easy adjustment of the aperture diaphragm, for bright field	Not meeting the specification
10	Ergo stage user selectable: user can select right or left hand operation. Stage plate with ultra hard ceramic surface, travel range 76mm x 25 mm, slide holder, X/Y stage control	Not meeting the specification

III. Due to administrative reason the tender for **KMSCL/EP/T498/756B//2023 : Advanced Fluorescent Microscope** is cancelled

**Dr. Shibulal A**  
General Manager