



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/4145/2024

Dated: 30.09.2024

NOTICE

Sub:- Tender No.T532/ 2024 - Cancelltion Notice – Neuro Musculator Stimulator TOF Monitor with Accelerometer - reg.

- Ref:-**
1. Tender Inviting Notice published dated 05.03.2024
 2. Technical bid opening held on 02.05.2024 & 13.05.2024 @ 11.00 A M
 3. Eligible list published on 20.07.2024
 4. Demonstration/ document evaluation of Neuro Musculator Stimulator TOF Monitor with Accelerometer conducted on 24.09.2024

I. The reports of technical document evaluation/ demonstration of Neuro Musculator Stimulator TOF Monitor with Accelerometer as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models
1.	Neuro Musculator Stimulator TOF Monitor with Accelerometer	M/s. SANMED HEALTH CARE	NMS450/ Xavant Technology

II. The other tenderers participated in the tenderer are rejected due to the reason mentioned against their name below;

A. M/s. Baxter (India) Private Limited

Rejected model;2510091TOF3D/ MIPM Mammendorfer

Reason for rejection:

Clause No	Technical specification	Deviations noted
6	Should have AUTO SMC mode, with Depth- of-block Zones to assess the depth of neuromuscular blockade	Not meeting the specification
8	Complete unit with Nerve Stimulator Cable (NMBA Cable) with TriAxialAccelerometer Sensor	Not meeting the specification

II. Due to administrative reason the tender for **KMSCL/EP/T532/1702/2024: Neuro Musculator Stimulator TOF Monitor with Accelerometer** is cancelled.

Dr. Shibulal A
General Manager

