



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking) Thycaud.P.O., Thiruvananthapuram-14. Tele Fax No. 0471-2945647 0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN: U24233KL2007SGC021616 GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/20600/2023

Dated: 09.04.2024

NOTICE

Sub:- Tender No.T519(R)/ 2023 - Results of demonstration/ document evaluation & Date of price bid opening – various equipment - reg.

Ref:- 1. Tender Inviting Notice published dated 09.08.2023

- 2. Technical bid opening held on 04.09.2023 @ 11.00 A M
- 3. Eligible list published on 20.10.2023
- 4. Reports of demonstration/ document evaluation of Ultrasonic Aspirator with Universal Hand Piece received on 02,04,2024
- 5. Demonstration/ document evaluation of various equipment conducted on 21.03.2024

I. The reports of technical document evaluation/ demonstration of various equipment as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

SI. No	Equipment	Name of the Selected Bidder	Selected Models
1.	Ultrasonic Aspirator with Universal Hand Piece	M/s Toshbro Medicals Pvt Ltd	Sonoca 300/Soering GmbH, Germany
2.	Patient Warmer	M/s Standard Medical Corporation	IOB-505/ IOB Medical Inc, USA
3.	DVT/ Sequential Compression Device Calf Pump	M/s Fortuna Technologies	SCD 700(Product Code 295259)/Cardinal Health

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

1. KMSCL/EP/T519(R)/165/2023: Patient Warmer

A. M/s Arkon Medical Systems

Rejected model: Therma STM-01/ Indian Surgical

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
11	Should have a safety certificate from a competent authority CE issued by a notified body registered in the European commission/FDA(US)/STQC CB certificate /STQC S certificate or a valid detailed electrica and functional safety test report from ERTL Copy of the certificate /test report shall be produced along with the technical bid	No valid certificate

2. KMSCL/EP/T519(R)/983/2023: DVT/ Sequential Compression Device Calf Pump

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
h	Controller should have Graphic user Interface of colour LCD screen which provides larger icons for greater visibility.	
n	Should have trouble shooting index in the device itself.	Not available

B. M/s Lakshmi Agency

Rejected model: DVT 7700/ Wonjin Mulsan Co, Ltd

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
	Should have trouble shooting index in the device	Not available
n	itself.	E.

C. M/s Universal Agencies

Rejected model; FLOWMIN 600/ MEDICOTRONICS

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
n	Should have trouble shooting index in the device itself.	Not available

III. The online price bid of the eligible tenderers mentioned above will be opened on v2.04.2024, @ 11.00 AM at the e- tender portal

Dr. Shibulal A

General Manager