



EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/662/2024

Dated: 26.06.2024

NOTICE

Sub:- Tender No.T522/ 2024 - Cancellation Notice – Laryngoscope Various Types - reg.

Ref:- 1. Tender Inviting Notice published dated 04.10.2023

2. Technical bid opening held on 27.10.2023 @ 11.00 A M

3. Eligible list published on 28.11.2024

4. Demonstration/ document evaluation of Laryngoscope Various Types conducted on 31.05.2024

I. The reports of technical document evaluation/ demonstration of Laryngoscope Various Types as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers participated in the technical evaluation are as follows:

1., M/s. Hansraj Nayyar Mediquip Inc

Rejected model: Anaesthetics India Pvt. Ltd

Reason for rejection: Sample not produced

2. M/s. Alan Medical & Laboratory Products

Rejected model: ALAN SCOPE/ALAN

Reason for rejection;

Clause No	Technical Specification	Deviations Noted
1	Should supply 4 different size standard blades and one handle for adult and pediatric separately and one short stubby handle	stubby handle not demonstrated

3. M/s. Kerala Surgical Equipment Co

Rejected model; Continental

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
2	Should be stainless Steel matt finished.	stubby handle not steel matt finished

4. M/s. ramon surgical company

Rejected model: F.3210-6, F.3201-6/SCOPE MEDICAL

Reason for rejection: Sample not produced

5. M/s. Trivandrum Surgicals and Lab Supplies

Rejected model: MAC/MILLER/ CHATTERJEE

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
2	Should be stainless Steel matt finished.	stubby handle not steel matt finished

6. M/s. UNIVERSAL AGENCIES..

Rejected model: SKU-15112 & SKU-15128/INDOSURGICAL PRIVATE LIMITED

Reason for rejection; Sample not produced

III. Due to technical reason the tender for **KMSCL/EP/T522/69/2023: Laryngoscope Various Types** is cancelled.

Dr. Shibulal A

General Manager