



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)
Thycaud.P.O., Thiruvananthapuram-14.
Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600 E-mail id : eptenders.kmscl@kerala.gov.in

CIN: U24233KL2007SGC021616 GSTIN: 32AADCK4029M1ZK

www.kmscl.kerala.gov.in

EP Division

No: KMSCL/EP/14383/2023 Dated: 02.12.2023

NOTICE

Sub: - Tender No.T502/ 2023 - Cancellation notice - Repetitive Transcranial Magnetic Stimulation Machine - reg.

Ref:- 1. Tender Inviting Notice published dated 09.08.2023

- 2. Technical bid opening held on 11.09.2023 & 15.09.2023 @ 11.00 A M
- 3. Eligible list published on 11.10.2023
- 4. Reports of demonstration/ document evaluation of Repetitive Transcranial Magnetic Stimulation Machine is received on 30.11.2023
- I. The reports of technical document evaluation/ demonstration of Repetitive Transcranial Magnetic Stimulation Machine as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

SI. N	Equipment	Name of the Selected Bidder	Selected Models
1.	Repetitive Transcranial Magnetic Stimulation Machine	M/s Aerobe Private Limited	MagPro R20+ Magverture A/S

II. The other tenderers participated in the tender are rejected due to the reason mentioned against their name below;

A. M/s Chroma N V

Rejected model: Neuro-MS/D/ Neurosoft LLC Russsia

Reason for rejection:

Clause No	Technical Specification	Compliance of the model			
2	Should be a portable single unit system suitable for Therapy & Research. Weight should be less than 25 kg.	No attached exterior			
8	Coil should be able to show the number of remaining stimuli and days on a screen inbuilt on the coil.	Not available			
11	Should have a facility for quick and simple protocol set- up on the stimulator.	Not available			
12	Should be supplied with an external control software to control the machine through Wi-Fi.	Not available			
12.b	Should be able to change following protocol parameters (repetition rate, pulses in train, number of trains, inter train interval, Start delay, Volume and motor threshold ratio) through Wi-Fi.	f Not available			
14	Software should be able to record messages via audio tab to be played before running the protocol.	Not available			
15	Should have inbuilt speakers so as to play audio to patients in regional language.	External system only available			

III.	Due	to	administrative	reasons	the	tender	for KMSCL/EP/T502/1676/2023	Repetitive	Transcranial	Magnetic
Stir	nulati	on l	Machine is cand	celled.						

Dr. Shibulal AGeneral Manager