



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/7831/2023

Dated: 08.12.2023

NOTICE

Sub:- Tender No.T492/ 2023 - Results of demonstration/ technical evaluation & Date of Price Bid Opening – various equipment - reg.

- Ref:-**
1. Tender Inviting Notice published dated 23.02.2023 & 04.05.2023
 2. Technical bid opening held on 13.03.2023, 18.03.2023, 24.03.2023, 27.03.2023 & 03.04.2023 @ 11.00 A M @ 11.00 A M
 3. Eligible list published on 21.08.2023, 20.09.2023
 4. Demonstration/ document evaluation of various equipment conducted on 29.11.2023

I. The reports of technical document evaluation/ demonstration of various items as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models
1.	Electronic Weighing Machine Lab -Type II	M/s Unique Scientific	CWS-302/Scaletec Mechatronics Pvt Ltd.
		M/s Universal Agencies	CG 302/ Aczet Pvt. Ltd.
		M/s Scientific Medical Equipments (Kerala) Pvt.	BI-220H/ Shimadzu (Asia Pacific)

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below:

A. KMSCL/EP/T492/1272/2023: Dry Bath

1. M/s. UNIVERSAL AGENCIES

Rejected model: RDI-SPL Controller/ B&C Instruments

Reasons for rejection: Sample not produced

2. M/s AAST Medi Tech Private Limited

Rejected model:DDI 24/AAST Medi Tech

Reason for rejection: Sample not produced

Clause No	Technical Specification	Deviations Noted
4	12mm, 16mm or 19mm test tubes or two block of 150x75x50 mm size for accommodating 100 Tubes of 10mm Diameter	One block can accommodate 36 tubes only