



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/673/2024

Dated: 16.04.2024

NOTICE

Sub:- Tender No.T523(R)/ 2023 - Results of demonstration/ technical evaluation & Date of Price Bid Opening – Advanced Fluorescent Microscope - reg.

- Ref:-**
1. Tender Inviting Notice published dated 16.01.2023
 2. Technical bid opening held on 05.02.2023 & 09.02.2024 @ 11.00 A M
 3. Eligible list published on 07.03.2024
 4. Reports of demonstration/ document evaluation OF Advanced Fluorescent Microscope received on 16.04.2024

I. The reports of technical document evaluation/ demonstration of Advanced Fluorescent Microscope as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models
1.	Advanced Fluorescent Microscope	M/s Bions Medical Systems Pvt Ltd	Lecia DM 2500 /Leica Microsystems

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

A. M/s Radical Scientific Equipments Pvt. Ltd

Rejected model: RXLr-5 (Code: NX/ Radical Scientific

Reason for rejection:

Clause No	Technical specification	Deviations noted
5	Should have 3- step focusing, coarse, medium and fine (1 and 4 microns)	Not meeting the specification
6	Five focusing functions (Coarse, medium, fine, focus height stops, adjustable torque)	Not meeting the specification
9	Condenser Achr.apl.A 0.9 (P) with switchable condenser head, with color coding for fast and easy adjustment of the aperture diaphragm, for bright field	Not meeting the specification
10	Ergo stage user selectable: user can select right or left hand operation. Stage plate with ultra hard ceramic surface, travel range 76mm x 25 mm, slide holder, X/Y stage control	Not meeting the specification

III. The online price bid of the eligible tenderer mentioned above will be opened on 19.04.2024 @ 11.00 am at the e-tender portal.

Dr. Shibulal A
General Manager