



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking) Thycaud.P.O., Thiruvananthapuram-14. Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600 E-mail id : <u>eptenders.kmscl@kerala.gov.in</u> CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

No: KMSCL/EP/20600/2023

www.kmscl.kerala.gov.in

Dated: 31.05.2024

NOTICE

Sub:- Tender No.T519(R)/ 2023 - Results of demonstration/ technical evaluation & Date of Price Bid Opening – various equipment - reg.

Ref:- 1. Tender Inviting Notice published dated 28.11.2023

- 2. Technical bid opening held on 11.12.2023 @ 11.00 A M
- 3. Eligible list published on 25.01.2024
- 4. Reports of demonstration/ document evaluation of Ophthalmology equipment received on 29.05.2024
- 5. Demonstration/ document evaluation of General surgery equipment conducted on 25.05.2024

I. The reports of technical document evaluation/ demonstration of various items as per reference no.4th & 5th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

SI. No	ltem	Name of the Selected Bidder	Model/OEM
1.	Laparoscopic System with Single Chip Camera		Karl Storz Laparoscopic System with, Single Chip Camera/ Karl Storz GmbH.
2.	SICS Set	M/s Kerala Surgical Equipment Co	Associated Surgicals
3.	Cataract Surgery Instruments	M/s Kerala Surgical Equipment Co.	Associated Surgical
4.	Autolensometer	M/s Remidio Innovative Solutions Private Limited	VX-36/ Luneau Technology
		M/s Lumen Mediworks Pvt Ltd	TL 6800/ NINGBO MINSING
5.	Ophthalmic Chair Unit	M/s Lumen Mediworks Pvt Ltd	MR2012/ TWOM OPHTHOTRONICS
		M/s Appasamy Associates Pvt Ltd	AARU 2000 Elegant/Appasamy

II. The tenderers participated in the tender are rejected due to the reasons mentioned against their name below:

A. KMSCL/EP/T519(R)/644/2023: Autolensometer

1. M/s Appasamy Associates Pvt Ltd.,

Rejected model: PLM8000/ Appasamy Associates

Reason for rejection:

Clause No	Technical specification	Deviations noted
6	Should have a wave length of 535 to 660nm	Not meeting the specification
8a	Should have spectacle lens size 20 to 108mm	Not meeting the specification
12	Should have UV tester, PD measuring, Auto progressive Lens measuring, distortion check	Distortion check not available.

III. The online price bid of the eligible tenderer mentioned above will be opened on U_{\uparrow} .05.2024 @ 11.00 am at the e-tender portal.

Dr. Shibulal AGeneral Manager