



**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [eptenders.kmscl@kerala.gov.in](mailto:eptenders.kmscl@kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

No: KMSCL/EP/14384/2023

Dated: 03.01.2024

**NOTICE**

**Sub:-** Tender No.T503/ 2023 - Results of demonstration/ technical evaluation & Date of Price Bid Opening – OT Light LED Type C - reg.

- Ref:-**
1. Tender Inviting Notice published dated 09.08.2023
  2. Technical bid opening held on 23.09.2023 @ 11.00 A M
  3. Eligible list published on 02.11.2023
  4. Reports of demonstration/ document evaluation of OT Light LED Type C is received on 28.12.2023

I. The reports of technical document evaluation/ demonstration of OT Light LED Type C as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models
1.	OT Light LED Type C	M/s Champak Enterprises Pvt Ltd	Q Flow/ MERIVAARA
		M/s Lakshmi Surgicals	Karl Storz LEDVISION101/ Karl Storz GmbH
		M/s Kerala Surgical Equipment Co	PSI-PENTA LUMEO/ Palakkad Surgical
		M/s ANTS Life Care Solutions LLP	SHALYA iLUX 20-20/ Xcellance Medical

II. The other tenderers participated in the tender are rejected due to the reason mentioned against their name below;

A. M/s Neoptics Surgicals

Rejected model: Chromophare F528-528/ Stryker

Reason for rejection: Different item demonstrated. Not Recommended

B. M/s Staan Bio-Med Engineering Private Limited

Rejected model: DISCOVER/ STAAN BIO-MED

Reason for rejection:

Clause No	Technical Specification	Compliance of the model
6.b	Variable Light patch diameter from 12 – 22cmsin each domes or better	Not meeting the specification
6.f	Depth of illumination (L1 + L2 ) of more than 50 cms	Not meeting the specification

C. M/s Midmark (India) Private Limited

Rejected model: Surgilight 4816/ Midmark

Reason for rejection:

Clause	Compliance of the
--------	-------------------

6.b	Variable Light patch diameter from 12 – 22cms in each domes or better	Not meeting the specification
6.f	Depth of illumination (L1 + L2 ) of more than 50 cms	Not meeting the specification

D. M/s HOSCO Private Limited

Rejected model: HOL-01/ HOSCO

Reason for rejection: Sample not produced.

III. The online price bid of the eligible tenderer mentioned above will be opened on 05.01.2024 @ 11.00 am at the e-tender portal.

**Dr. Shibulal A**  
General Manager