



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/4227/2024

Dated: 12.09.2024

NOTICE

Sub:- Tender No.T538(R)/ 2024 - Results of demonstration/ technical evaluation & Date of Price Bid Opening – Electrolyte Analyzer - reg.

Ref:- 1. Tender Inviting Notice published dated 05.03.2024

2. Technical bid opening held on 05.04.2024 & 15.04.2024 @ 11.00 A M

3. Eligible list published on 08.08.2024

4. Demonstration/ document evaluation of Electrolyte Analyzer conducted on 27.08.2024

I. The reports of technical document evaluation/ demonstration of Electrolyte Analyzer as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models
1.	Electrolyte Analyzer	M/s. Transasia Bio Medicals Ltd	EC 90/ Erba Lachema
		M/s. Sensa Core Medical Instrumentation Pvt Ltd	ST-200 aQua
		M/s. UNIVERSAL AGENCIES	AGD3100/AGD BIOMEDICALS
		M/s. LabX	LabX/Qline Biotech Pvt Ltd
		M/s. New India Surgicals and Chemicals	Quantilyte (NA//K/CL)/ Erba Lachema

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

A. M/s. DiaSys Diagnostics India Private Limited

Rejected model : EasyLyte/ Medica Corporation

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
1	Should be able to measure sodium potassium, ionized calcium & chloride.	The demonstrated machine has no provision for measuring ionized calcium.

III. The online price bid of the eligible tenderer mentioned above will be opened on 18.09.2024 @ 11.00 am at the e-tender portal.