



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/4145/2024

Dated: 25.09.2024

NOTICE

Sub:- Tender No.T532/ 2024 -Cancellation Notice– Fully Manual Rotary Microtome - reg.

- Ref:-**
1. Tender Inviting Notice published dated 05.03.2024
 2. Technical bid opening held on 02.05.2024 & 13.05.2024 @ 11.00 A M
 3. Eligible list published on 20.07.2024
 4. Demonstration/ document evaluation of Fully Manual Rotary Microtome conducted on 13.09.2024

I. The reports of technical document evaluation/ demonstration of Fully Manual Rotary

Microtome as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

| Sl. No | Equipment | Name of the Selected Bidder | Selected Models |
|--------|-------------------------------|------------------------------------|--|
| 1. | Fully Manual Rotary Microtome | M/s. Bions Medical Systems Pvt Ltd | Leica RM2125 RTS Rotary Microtome/Lecia Biosystems |

II. The other tenderers participated in the tender are rejected due to the reason mentioned against their name below;

1. M/s. UNIVERSAL AGENCIES

Rejected model: CUT 4062/ SLEE MEDICAL

Reason for rejection:

| Clause No | Technical specification | Deviations noted |
|-----------|--|-------------------------------|
| 8 | Mechanical trimming thickness selection of 10 and 50 µm. Optional selection of 20,30 and 40 µm | Not meeting the specification |
| 15 | Additional 12 O' Clock locking by simply pulling the handle | Not convinced |

2. M/s. Vision Diagnostic India Pvt. Ltd

Rejected model:HM325/ Epredia

Reason for rejection:

| Clause No | Technical specification | Deviations noted |
|------------------|--|-------------------------------|
| 15 | Additional 12 O' Clock locking by simply pulling the handle | Not meeting the specification |
| 17 | Coarse feed wheel turn direction: User selectable | Not meeting the specification |
| 18 | The instrument can be used optionally with either clockwise rotation or counter clockwise rotation for the coarse feed wheel | Not meeting the specification |

III. Due to administrative reason the tender **KMSCL/EP/T532/626/2024: Fully Manual Rotary Microtome** gets cancelled.

Dr. Shibulal A
General Manager