



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)
Thycaud.P.O., Thiruvananthapuram-14.
Tele Fax No. 0471-2945647
0471-2945646, 0471-2945600

E-mail id: eptenders.kmscl@kerala.gov.in CIN: U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

No: KMSCL/EP/662/2024

www.kmscl.kerala.gov.in

Dated: 04.04.2024

NOTICE

- Sub:- Tender No.T522/ 2024 Results of demonstration/ technical evaluation & Date of Price Bid Opening various equipment reg.
- Ref:- 1. Tender Inviting Notice published dated 04.10.2023
 - 2. Technical bid opening held on 27.10.2023 @ 11.00 A M
 - 3. Eligible list published on 28.11.2024
 - 4. Demonstration/ document evaluation of various equipment conducted on 22.03.2024 & 23.03.2024
- I. The reports of technical document evaluation/ demonstration of various equipment as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

SI No	Equipment Name	Name of the Selected Bidder	Model/OEM
1.	Emergency Trolley (Trauma Emergency Cot)	M/s Carevel Medical System Pvt Ltd	C-2002-H/ Carevel Medical
		M/s Imedfurns Pvt Ltd	IMED5602K/ IMEDFURNS
		M/s Medimek Industries	Mi-4076 A/ Medimek Industries
2.		M/s Carevel Medical System	C-4201-A/ Carevel
		M/s Imedfurns Pvt Ltd	IMED5906K/ IMEDFURNS
		M/s Kerala Surgical Equipment Co.	MT-4/ Medik Enterprises
		M/s Medimek Industries	Mi-5036 X/ Medimek
3.	Bowl with Stand	M/s Dynamic International	DHWE-1122/ DYNAMIC INTERNATIONAL
		M/s Kerala Surgical Equipment Co	MT-BS/ Medik Enterprises
4.	Linen Trolley	M/s Dynamic International	DHF-2317/ DYNAMIC INTERNATIONAL
		M/s Imedfurns Pvt Ltd	IMED6014K/ IMEDFURNS
		M/s Kerala Surgical Equipment Co	MT-LT100/ Medik Enterprises

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

A KMSCL/EP/T522/551Z/2024: Emergency Trolley (Trauma Emergency Cot)

1. M/s Alan Medical & Laboratory Products

Rejected model: PTT-03/ MEDI WAVES INC Reason for rejection: Sample not produced

Rejected model: DHPT-1150/ DYNAMIC INTERNATIONAL

Reason for rejection:

2. M/s Dynamic International

Clause No	Technical Specification	Deviations Noted	
1	Should have two sectional top Xray translucen thigh pressure laminate with Cutting of the sheet not facility to insert Xray cassette from both proper with sharp edges sides and ends of trolley		
2	Should be able to X ray the patient from positions the entire length and width of the trolley	X-Ray cassette tray is	
5	Should have hydraulic height adjustment.	Noise during height adjustment and down movement not working without any load on the stretcher top	
8	Should have detachable SS telescopic IV rods and collapsible SS side rails.	End of the telescopic IV rods are plastic	
11.iii	Height:Min 65±5cm Max90±5cm	not meeting the specification	

B. KMSCL/EP/T522/551I/2024: Medicine Trolley

1. M/s Dynamic International

Rejected model; DHF-1190/ DYNAMIC

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
	Comments	Aesthetics of the welded joints are not good and poor workmanship

III. The online price bid of the eligible tenderer mentioned above will be opened on %.03.2024 @ 11.00 am at the e-tender portal.