



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/4146/2024

Dated: 26.09.2024

NOTICE

Sub:- Tender No.T533/ 2024 - Results of demonstration/ document evaluation & Date of price bid opening – Leg Stirrups with Lift Assist Technology - reg.

- Ref:-**
1. Tender Inviting Notice published dated 01.02.2024
 2. Technical bid opening held on 17.02.2024 @ 11.00 A M
 3. Eligible List published on 19.06.2024
 4. Reports of demonstration/ document evaluation of Leg Stirrups with Lift Assist Technology received on 23.09.2024

I. The reports of technical document evaluation/ demonstration of Leg Stirrups with Lift Assist Technology as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

SI No	Equipment Name	Selected Bidders	Model/OEM
1.	Leg Stirrups with Lift Assist Technology	M/s. BET Medical Private Ltd	KYRA3000/Kyra Medical

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

A. M/s. CHAMPAK ENTERPRISES PVT LTD

Rejected model:Magna - Fins/Magna - Fins

Reason for rejection:

Clause no	Technical specification	Compliance of the model
11	Boot should be self adjusting type to minimize pressure on the calf	Not satisfactory
	Comments	The universal clamp for fixing the stirrups was found NOT SUITABLE for the OT table at the time of demonstration.

B. M/s. Sree Gokulam Healthcare Pvt. Ltd.

Rejected model: A62XX Standard Stirrup HyStirrup III/ Nanjing Mindary

Reason for rejection:

Clause no	Technical specification	Compliance of the model
8	Indicators/ scales marking should be available for precise positioning	No indicator for LEG holder movements.
13	Should supply specialized trolley to hang the stirrups when not in use	Trolley facility not available

Comments	stirrups was found NOT SUITABLE for the OT table at the time of demonstration.
----------	--

III. the final L1 bidder is requested to submit final negotiated offer for the Unit cost , AMC, CMC with similar supply order copy on or before 30.09.2024

Dr. Shibulal A
General Manager