



**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [eptenders.kmscl@kerala.gov.in](mailto:eptenders.kmscl@kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

No: KMSCL/EP/17267/2023

Dated: 03.01.2024

**NOTICE**

**Sub:-** Tender No.T512/ 2023 - Results of demonstration/ technical evaluation & Date of Price Bid Opening – various equipment - reg.

- Ref:-**
1. Tender Inviting Notice published dated 04.10.2023
  2. Technical bid opening held on 30.10.2023 @ 11.00 A M
  3. Eligible list published on 03.11.2023
  4. Demonstration/ document evaluation of Micro pipette various types & Centrifuge- various Types conducted on 14.12.2023
  5. Demonstration/ document evaluation of Lead Apron & Token Display conducted on 16.12.2023

I. The reports of technical document evaluation/ demonstration of various equipment as per reference no. 4th & 5th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models
1.	Micro pipette (variable volume)	M/s BIOGENE	Transferpette@α, Transferpette@s/BRAND Scientific
2.	Centrifuge- various Types	M/s Scientific Medical Equipments (Kerala) Pvt. Ltd	TC 650F(Centrifuge)TC 652 (Rotor),TC 650F (Centrifuge) TC 654 (Rotor)/ELEKTROCRAFT
		M/s Universal Agencies	R 8C PLUS with S 16-15ml Rotor, MEDICO Pluse with SM 8-15ml Rotor/ REMI
3.	Lead Apron	M/s Alan Medical & Laboratory Products	Lead Lite/A.P. Enterprices
		M/s C Cube Advanced Technologies	AT-100/ AMRAD MEDICAL
		M/s Trivandrum Surgicals and Lab Supplies	3204LL-X CORDON FRONTAL LEAD APRON& 3213LL-X THYROID GOARD/ HORIZON
4.	Token Display Units	M/s Data Care System Pvt.Ltd	TDCM, TKCM, SINGLE LINE TOKEN DISPLAY, KJS40ASFHD, TL-WR841N/ DATA CARE
		M/s Biocare Diagnostics	TDS003/ INVADE EMBEDDED

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

**a. Micro pipette (variable volume)**

A. M/s Universal Agencies

Rejected model/ OEM: SLP SERIES (VARIABLE), SLP SERIES (FIXED)/MICROLIT

Reason for rejection:

Clause no	Technical specifications	Deviations
3	Control button shall be operated by very low operating force.	(upto 920 gm force needed)
	Comments	Unacceptable variation in precision and accuracy at lower volumes.

B. M/s Scientific Medical Equipments (Kerala) Pvt

Rejected model: Vertex Premium Variable Volume, Vertex Premium Fixed Volume/RV INSTRUMENTS

Reason for rejection:

Clause no	Technical specifications	Deviations
1	Should have 4 digits volume display.	Required volume ranges were not given for evaluation
2	Should be a fully autoclavable type.	
3	Control button shall be operated by very low operating force.	
4	Ejector shall be operated by very low operating force.	
5	Should be resistant to UV light.	
6	All available volume ranges should be mentioned in the offer form.	
7	Should have volume ranges of 10-200 µl, 200 -1000µl, 20 - 200µl, 100 - 1000µl, 5 - 10 µl, 5 - 50 µl, 100 -500 µl, 0-2 – 2 µl, 1-10 µl, 2 -20 µl, 30 -300 µl, (The bidder can offer any other volumes available in the offer form)	
8	Pipette should be IVD certified.	

**b. Micro pipette (Fixed volume)**

A. M/s BIOGENE

Rejected model: Transferpette@α, Transferpette@s/BRAND Scientific

Reason for rejection:

Clause no	Technical specifications	Deviations
1	Should be a fully autoclavable type.	Required volume ranges were not given for evaluation
2	Control button shall be operated by very low operating force.	
3	Ejector shall be operated by very low operating force.	
4	Should be resistant to UV light.	
5	Pipette should be IVD certified.	
	All available volumes should be mentioned	were not given for evaluation

	in the offer form.	
7	Should have volume ranges of 1 µl, 10 µl, 20 µl, 50 µl, 100 µl, 200 µl, 25 µl, 250 µl, 500 µl, 1000 µl, 5000 µl, 10000 µl, (The bidder can offer any other volumes available in the offer form)	

B. M/s Universal Agencies

Rejected model: SLP SERIES (VARIABLE), SLP SERIES (FIXED)/MICROLIT

Reason for rejection:

Clause no	Technical specifications	Deviations
1	Should be a fully autoclavable type.	Required volume ranges were not given for evaluation
2	Control button shall be operated by very low operating force.	
3	Ejector shall be operated by very low operating force.	
4	Should be resistant to UV light.	
5	Pipette should be IVD certified.	
6	All available volumes should be mentioned in the offer form.	
7	Should have volume ranges of 1 µl, 10 µl, 20 µl, 50 µl, 100 µl, 200 µl, 25 µl, 250 µl, 500 µl, 1000 µl, 5000 µl, 10000 µl, (The bidder can offer any other volumes available in the offer form)	

C. M/s Scientific Medical Equipments (Kerala) Pvt Ltd

Rejected Model: Vertex Premium Variable Volume, Vertex Premium Fixed Volume/RV INSTRUMENTS

Reason for rejection:

Clause no	Technical specifications	Deviations
1	Should be a fully autoclavable type.	Required volume ranges were not given for evaluation
2	Control button shall be operated by very low operating force.	
3	Ejector shall be operated by very low operating force.	
4	Should be resistant to UV light.	
5	Pipette should be IVD certified.	
6	All available volumes should be mentioned in the offer form.	
7	Should have volume ranges of 1 µl, 10 µl, 20 µl, 50 µl, 100 µl, 200 µl, 25 µl, 250 µl, 500 µl, 1000 µl, 5000 µl, 10000 µl, (The bidder can offer any other volumes available in the offer form)	

2. KMSCL/EP/T512/71/2023: Lead Apron

A. M/s X Ray Medico Surgical Engineering Ancillary

Reason for rejection:

Clause no	Technical specifications	Deviations
1	Should be AERB approved.	AERB certificate not produced.

III. Due to administrative reason the tender for **KMSCL/EP/T512/180/2023: Micro Pipette - Various Types** is cancelled.

IV. The online price bid of the eligible tenderer mentioned above will be opened on 08.01.2024 @ 11.00 am at the e-tender portal.

**Dr. Shibulal A**  
General Manager