



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/10408/2024

Dated: 30.09.2024

NOTICE

Sub:- Tender No.T544(R)/ 2024 - Results of demonstration/ technical evaluation & Date of Price Bid Opening – Biological Effluent Decontamination Plant - reg.

- Ref:-**
1. Tender Inviting Notice published dated 16.01.2024
 2. Technical bid opening held on 31.01.2024, 05.02.2024 @ 11.00 A M
 3. Eligible list published on 19.06.2024
 4. Reports of demonstration/ document evaluation of Biological Effluent Decontamination Plant received on 27.09.2024

I. The reports of technical document evaluation/ demonstration of Biological Effluent Decontamination Plant as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

SI No	Equipment	Name of the Selected Bidder	Selected Models
1.	Biological Effluent Decontamination Plant	M/s. Klenzaid's Contamination Controls Private Limited	Klenzaid's/Klenzaid's Contamination

II. The other tenderers participated in the tender are rejected due to the reason mentioned against their name below;

1. M/s. VA Exergy Technologies India Pvt td
Rejected model: ETP/VA ExergyTecnologies
Reason for rejection:

Clause No	Technical Specification	Deviations Noted
III. 7	All the tank and connectors should be high quality steel with anticorrosive property, SS 304 or higher grade.	No,Boiler based & boiler is MS
V. 10	The details of installation of the unit in BSL2 OR BSL3 lab should be mentioned.	Only one installation of BETP in JIPMER

III. The online price bid of the eligible tenderer mentioned above will be opened on 0310.2024 @ 11.00 am at the e-tender portal.

Dr. Shibulal A
General Manager