

**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : etenders_ep@kmscl.kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

Equipment Divisionwww.kmscl.kerala.gov.in**No: KMSCL/EP/1950/2021****Dated: 10.08.2021****NOTICE**

Sub:- Tender No.T391/2021 - Results of representation received against the technical document evaluation/ demonstration & Date of Price Bid Opening – Hematology Analyzer 3 Part Model A - reg.

- Ref:-**
1. Tender Inviting Notice published dated 11.02.2020
 2. Technical bid opening held on 10.03.2021,22-03-2021 @ 11.00 A M
 3. Eligible list for demonstration published on 04.05.2021
 4. Results of demonstration/ document evaluation of HEMATOLOGY ANALYZER 3 PART MODEL A notice on 05.08.2021.
 5. Representation email received on 09.08.2021
 6. Reports of evaluation of representation dated 10.08.2021

I. With reference to the demonstration result notice published for Hematology Analyzer 3 Part Model A vide 4th above a representation was submitted by the rejected tenderer - M/s. Triviron Healthcare Pvt. Ltd vide 5th above. The representation received was evaluated in detail and the results are as follows;

The technical committee evaluated the representation and the results print outs of the tenderer - Triviron Healthcare Pvt. Ltd again. It was found that the RBC value used for calculation of CV% is wrongly entered – instead of 2.27 it is used 2.77. Based on the correct value that is 2.27 the CV% is 0.24% and it is in the acceptable limit as per the tender stipulated condition (3 level QC should be demonstrated to the technical committee during the technical demonstration of the offered equipment. All 3 levels of QC should be run in specified time and in sequence. Any 3rd party QC should be used and should be run in QC mode. The sequence of QC run (all 3 levels) will be intimated during demo. The CV of RBC, WBC, HB and PLT should be less than or equal to 5% and Low PLT QC should be less than or equal to 10%).

Parameters	L1 - 1	L1 - 2	L1 - 3	L1 - 4	L1 - 5	CV - L1
WBC	1.90	1.90	1.90	1.90	1.90	0.00
RBC	2.26	2.27	2.27	2.26	2.26	0.24
HB	6.00	6.00	6.00	6.00	6.00	0.00
PLT	70.00	69.00	63.00	65.00	62.00	5.42

Hence the representation is considered and the offer of M/s Triviron Healthcare Pvt Ltd is selected for price bid opening.

II. The eligible list for price bid opening is revised as follows;

Sl. No	Equipment	Name of the Selected Bidder	Selected Model
1.	Hematology Analyzer 3 Part Model A	M/s Deep Meditech Private Limited	BC-21 S/ Shenzhen Mindray
		M/s. AGD Biomedicals Pvt Ltd	AGD 300/ Genrui Biotech
		M/s. LabX	M 32 B/ Boule Medical
		M/s. Transasia Bio Medicals Ltd	H-360/ Erba
		M/s. Triviron Healthcare Pvt. Ltd	Cellenium 21/ Triviron

II. The online price bid of the eligible tenderers mentioned above will be opened on 11.08.2021 @ 03.00 pm at the e-tender portal.

Dr. Dileep Kumar S R
General Manager