



**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [etenders\\_ep@kmscl.kerala.gov.in](mailto:etenders_ep@kmscl.kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

No: KMSCL/EP/4057/2022

Dated: 22.06.2022

**NOTICE**

**Sub:-** Tender No.T436/ 2022 - Results of demonstration/ technical evaluation & Date of Price Bid Opening –  
Holter Monitor Model B- reg.

- Ref:-**
1. Tender Inviting Notice published dated 03.03.2022
  2. Technical bid opening held on 08.04.2022 @ 11.00 A M
  3. Reports of demonstration/ document evaluation of Holter Monitor Model B received on 17.06.2022

I. The reports of technical document evaluation/ demonstration of Holter Monitor Model B as per reference no. 3rd, cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Item	Name of the Selected Bidder	Selected Model
1.	Holter Monitor Model B	M/s. UNIVERSAL AGENCIES	H12B PLUS/SHENZHEN
		M/s. Gemini Medical Networks Pvt Ltd.	DMS 300-4AL/DM Systems
		M/s. Schiller Healthcare India Pvt.Ltd..	Medilog AR/Schiller AG

II. The other tenderers participated in the tender are rejected due to the reasons mention against their name below;

A. M/s. RECORDERS AND MEDICARE SYSTEMS PVT LTD.

Rejected Model; CORDIS 12/Recorders and Medicare

Reason for rejection:

Clause No	Technical specification	Deviations noted
5.	Should analysis ST segments with ST episode reporting and heart rate variability on time and frequency domain.	ST analysis not good.
9.	Trend graphs- HR, RR interval, ST should be possible.	No trend graph for ST

III. The online price bid of the eligible tenderer mentioned above will be opened on 24.06.2022 @ 11.00 am at the e-tender portal.

**Dr. S S Joy**  
General Manager