



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/14386/2023

Dated: 30.01.2024

NOTICE

Sub:- Tender No.T505(R)/ 2023 - Cancellation Notice – Dental X- Ray Machine - reg.

Ref:- 1. Tender Inviting Notice published dated 09.08.2023

2. Technical bid opening held on 13.09.2023 @ 11.00 A M

3. Eligible list published on 28.11.2023

4. Demonstration/ document evaluation of Dental X- Ray Machine conducted on 18.01.2024

I. The reports of technical document evaluation/ demonstration of Dental X- Ray Machine as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers participated in the technical evaluation are as follows:

1. M/s Iatome Electric India Pvt Ltd

Rejected model: Aleio DC Optima/ IATOME ELECTRIC

Reasons for rejection:

Clause No	Technical Specification	Deviations Noted
2	Focal Spot 0.7mm or better	Focal Spot size 0.8 mm
3	Tube voltage Selectable 60-70KVp or better	Non selectable – 65 KVp fixed
4	Current Selectable 4, 7mA or better	Non selectable – 7mA fixed.
5	Exposure time 0.04-2 Sec	0.01 sec to 1 sec

III. Due to administrative reason the tender for **KMSCL/EP/T505(R)/31/2023: Dental X- Ray Machine** is cancelled.

Dr. Shibulal A

General Manager