



EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/20599/2023

Dated: 08.04.2024

NOTICE

Sub:- Tender No.T518/ 2023 - Results of demonstration/ technical evaluation & Date of Price Bid Opening – various equipment - reg.

- Ref:-**
1. Tender Inviting Notice published dated 28.11.2023
 2. Technical bid opening held on 15.12.2023 @ 11.00 A M
 3. Eligible list published on 07.03.2024
 4. Demonstration/ document evaluation of various items conducted on 20.03.2024 & 21.03.2024

I. The reports of technical document evaluation/ demonstration of various equipment as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models
1.	Multiparameter Monitor Neonatal	M/s Arkon Medical Systems	STAR8000F/ Shenzhen Comen Medical
		M/s Sree Gokulam Healthcare Pvt. Ltd	uMEC7/ Shenzhen Mindray
		M/s Universal Agencies	XM750/ Shenzhen Hwatime
2.	Fumigation dispenser	M/s Kerala Surgical Equipment Co	Aerojet Steriz/ Radiant Enterprises

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below:

1. KMSCL/EP/T518/50/2023: Fumigation dispenser

A. M/s Lakshmi Agency

Rejected model: 3019 PD/ Idealin fogging system

Reason for rejection: Sample not produced

2. KMSCL/EP/T518/105B/2023: Pulse Oxymeter Baby And Adult

A. M/s Arkon Medical Systems

Rejected model: Accuro/ Charmcare Co.Ltd

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
4	Should provide bar graph for pulse strength. Should have low perfusion and motion tolerance technology (510k Summary)	Not submitted

B. M/s Mann Electronics India Pvt. Ltd

Rejected model : MN-1014 C1/ MANN ELECTRONICS

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
4	Should provide bar graph for pulse strength. Should have low perfusion and motion tolerance technology (510k Summary)	Not submitted
12	Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid	No valid certificate

C. M/s MITI Electronics Pvt.Ltd

Rejected model: BT 720/ BISTOS

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
4	Should provide bar graph for pulse strength. Should have low perfusion and motion tolerance technology (510k Summary)	Not submitted

III. Due to administrative reason the tender for **KMSCL/EP/T518/50/2023 : Fumigation dispenser & KMSCL/EP/T518/105B/2023: Pulse Oxymeter Baby And Adult** are cancelled.

IV. The online price bid of the eligible tenderer mentioned above will be opened on **12.04.2024 @ 11.00 am** at the e-tender portal.

Dr. Shibulal A
General Manager