



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/11989/2023

Dated: 18.12.2023

NOTICE

Sub:- Tender No.T500/ 2023 - Results of demonstration/ technical evaluation & Date of Price Bid Opening – OT Integration - reg.

- Ref:-**
1. Tender Inviting Notice published dated 05.07.2023
 2. Technical bid opening held on 04.09.2023 @ 11.00 A M
 3. Eligible list published on 11.10.2023
 4. Reports of demonstration/ document evaluation of OT Integration received on 15.12.2023

I. The reports of technical document evaluation/ demonstration of OT Integration as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models
1.	OT Integration	M/s Lakshmi Surgicals	Karl Storz/ Karl Storz

II. The other tenderers participated in the tender is rejected due to the reasons mentioned against their name below:

A. M/s Champak Enterprises PVT LTD

Rejected Model/OEM: ADVANTECH

Reason for rejection:

1. The offered items are not from a single Original Equipment Manufacturer. The Video Image Management & communication system, streaming and audio components are from Kramer, Graphic user interface, digital documentation system are from Advantech, PTZ room camera is from Sony and the On-Air lamp is offered as local solution.

As per tender clause 6.1.1 "Manufacturers or their authorized dealers/Indian subsidiaries/direct importers having a place of business in any of the States of India are eligible to participate in this tender". In the tender manufacturers authorisation form of two different companies Advantech and Kramer is submitted. There cannot be two manufacturers for one product. The solution provided by the bidder is from multiple manufacturers, which is not complying with tender conditions.

2. Documents to substantiate "System should be SE-46 whitelisted and have AES-256 Database encryption for highest standard of cyber security" is not submitted.

3. As per the user list provided by the bidder none of the 12 users has acknowledged using bi-directional telestration in their respective hospital.

4. During technical evaluation the representative informed the committee that the offered system is not 3D compatible. Later he said it is compatible to 3D. However, on telephonic discussion with the user details provided by the bidder, none of the user has acknowledged usage of 4K 3D in their hospitals.

5. As per technical specification clause I, The system should be medical grade and provided for each Operation theatre. Documents to substantiate the offered product is medical grade is not submitted.

III. The online price bid of the eligible tenderer mentioned above will be opened on 21/12.2023 @ 11.00 am at the e-tender portal.

Dr. Shibulal A
General Manager