



EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/1135/2024

Dated: 09.04.2024

NOTICE

Sub:- Tender No.T524(R)/ 2023 - Results of demonstration/ technical evaluation & Date of Price Bid Opening – Decontamination Autoclave - reg.

- Ref:-**
1. Tender Inviting Notice published dated 04.10.2023
 2. Technical bid opening held on 27.10.2023 @ 11.00 A M
 3. Eligible list published on 28.11.2024
 4. Reports of demonstration/ document evaluation of Decontamination Autoclave conducted on 01.04.2024

I. The reports of technical document evaluation/ demonstration of Decontamination Autoclave as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models
1.	Decontamination Autoclave	M/s HICONTAINMENTS INTERNATIONAL PVT LTD	450x450x900mm/M/s Gemerik

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

A. M/s Genist Technocracy Pvt Ltd.

Rejected model:GTA-006/ GENIST

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
	Comments	Rejected due to non availability of service center facilities in Kerala/ South India

B. M/s Kerala Surgical Equipment Co

Rejected model: Medipharm

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
V.2.f	Decontamination autoclave must be connected to the building management system (BMS) for efficient control and monitoring of the system.	Not meeting the specification
VIII.3	The sterilizer should have either RS232 or Ethernet port to facilitate connectivity for network applications.	Not meeting the specification
XIX .1	The product should meet the following provisions and standards: Europe EN285 for large autoclave.	Not submitted

	Certificate.	
XIX .3	Pressure equipment directives: PED 97/23 EC.	Not submitted
XIX .4	ISO 13485:2003 Quality system for medical devices.	Not submitted

C. M/s Zenotrex Medico Pvt.Ltd

Rejected model:ZPL-AH02F/ Zenotrex Medico

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
	Comments	Rejected due to non availability of service center facilities in Kerala/ South India

12.04.2024

III. The online price bid of the eligible tenderer mentioned above will be opened on @ 11.00 am at the e-tender portal.

Dr. Shibulal A
General Manager