



**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [eptenders.kmscl@kerala.gov.in](mailto:eptenders.kmscl@kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

**EP Division**

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

**No: KMSCL/EP/11990/2023**

**Dated: 13.03.2024**

**NOTICE**

**Sub:-** Tender No.T501(R)/ 2023 - Cancellation Notice – Salters Baby Weighing Machine - reg.

- Ref:-**
1. Tender Inviting Notice published dated 05.07.2023
  2. Technical bid opening held on 31.07.2023 @ 11.00 A M
  3. Eligible list published on 20.09.2023
  4. Reports of demonstration/ document evaluation Salters Baby Weighing Machine received on 12.02.2024 & 16.02.2024

I. The reports of technical document evaluation/ demonstration of Salters Baby Weighing Machine as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers participated in the technical evaluation are as follows:

A. M/s Kerala Surgical Equipment Co  
 Rejected model; Doctor Beliram & Sons  
 Reason for rejection:

Clause No	Technical Specification	Deviations Noted
5	Should be supplied with top and bottom hook with weighing trouser	weighing trouser is too small with bearing capacity of 5 kg

B. M/s Marc Weighing Systems  
 Rejected model: 235-6M & 6S/DOCTOR BELI RAM  
 Reason for rejection: Sample not produced.

II. Due to administrative reason the tender for **KMSCL/EP/T501(R)/145B/2022: Salters Baby Weighing Machine** is cancelled.

**Dr. Shibulal A**  
 General Manager