



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)
Thycaud.P.O., Thiruvananthapuram-14.
Tele Fax No. 0471-2945647
0471-2945646, 0471-2945600

E-mail id: eptenders.kmscl@kerala.gov.in
CIN: U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

www.kmscl.kerala.gov.in

Dated: 05.10.2024

EP Division

No: KMSCL/EP/19056/2023

NOTICE

Sub:- Tender No.T517(R)/ 2023 - Results of demonstration/ technical evaluation & Date of Price Bid Opening – various equipment - reg.

Ref:- 1. Tender Inviting Notice published dated 03.11.2023

- 2. Technical bid opening held on 27.11.2023 & 23.12.2023 @ 11.00 A M
- 3. Eligible List published on 25.01.2024
- 4. Reports of demonstration/ document evaluation of various equipment conducted on received on 01.10.2024

I. The reports of technical document evaluation/ demonstration of various equipment as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

SI No	Equipment Name	Selected Bidders	Model/OEM
1.	count 2000	M/s. Carestream Health India Pvt	
2.	Type B Digital Radiographic System Type C	M/s. Carestream Health India Pvt Ltd M/s.Allengers Medical Systems	Health MARS-50/ Allengers
3	Digital Radiographic System Type D	M/s.Allengers Medical Systems	medical MARS-30/ Allengers medical

II The other tenderer participated in the tender are rejected due to the reasons mentioned against their name below:

1. KMSCL/EP/T517(R)/34 B/2023 : Digital Radiographic System Type B

A. M/s. Trivitron Healthcare Pvt. Ltd.

Rejected model: Ultisys 82/ Kiran Medical Systems

Reason for rejection:

Clause No	Technical Specification	Deviations Noted	
V.1	Should be a carbon fibre/equivalent motorized fixed 6 / 4 way movement floating table having a weight carrying capacity of 200kgs or more		
VI.1	Should be capable to take images in horizontal vertical and oblique positions with suitable motorized movements for all skeletal body including spine and chest	oblique not demonstrated.	
VIII.2	It should have the possibility of acquiring the image from the detector system. Should have not demonstrated		

2. KMSCL/EP/T517(R)/34 C /2023 : Digital Radiographic System Type C

A. M/s.Trivitron Healthcare Pvt. Ltd.

Rejected model: Ultisys 52/ Kiran Medical Systems

Reason for rejection:

Clause No	Technical Specification	Deviations Noted	
V.1	Should be a carbon fibre/equivalent motorized fixed 6 / 4 way movement floating table having a weight carrying capacity of 200kgs or more	Only vertical movement motorized	
VI.1	Should be capable to take images in horizontal vertical and oblique positions with suitable motorized movements for all skeletal body including spine and chest	oblique not demonstrated.	
VIII.2	It should have the possibility of acquiring the image from the detector system. Should have not demonstrated preview time 5 seconds or better		
	Comments	Perfomance certificates has no name & personal seals of HoD's	

3. KMSCL/EP/T517(R)/34 D/2023: Digital Radiographic System Type D

A. M/s.Trivitron Healthcare Pvt. Ltd.

Rejected model: Ultisys 40/ Kiran Medical Systems

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
V.1	Should be a carbon fibre/equivalent motorized fixed 6 / 4 way movement floating table having a weight carrying capacity of 200kgs or more	Only vertical movement motorized
VI.1	Should be capable to take images in horizontal, vertical and oblique positions with suitable motorized movements for all skeletal body including spine and chest	
VIII.2	It should have the possibility of acquiring the image from the detector system. Should have not demonstrated preview time 5 seconds or better	
	Comments	Performance certificates has no name & personal seals of HoD's

III. The online price bid of the eligible tenderer mentioned above will be opened on 08.10.2024 @ 11.00 am at the e-tender portal.