



**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [eptenders.kmscl@kerala.gov.in](mailto:eptenders.kmscl@kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

**EP Division**

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

**No: KMSCL/EP/17266/2023**

**Dated: 28.02.2024**

**NOTICE**

**Sub:-** Tender No.T511/ 2023 - Cancellation Notice – Fully Loaded Dental Chair Electrically Operated - reg.

- Ref:-**
1. Tender Inviting Notice published dated 04.10.2023
  2. Technical bid opening held on 06.11.2023 , 23.11.2023 @ 11.00 A M
  3. Eligible list published on 25.01.2024
  4. Reports of demonstration/ document evaluation of Fully Loaded Dental Chair Electrically Operated received on 08.02.2024

I. The reports of technical document evaluation/ demonstration of Fully Loaded Dental Chair Electrically Operated as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models
1.	Fully Loaded Dental Chair Electrically Operated	M/s Confident Dental Equipment Ltd.	Mookambika Dental Chair

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

A. M/s Trichur Surgicals

Rejected model/ OEM: TS 431/ TRISCO

Reason for rejection:

Clause No	Technical Specification	Compliance of the model
2.	Two Airtor points with tubing & metal nut fittings with Airtor standard head hand piece 1 No.(NSK or equivalent)	Inferior quality chuck type hand piece with standard head
8.	High quality water booster	Made of powder coated metal
14.	Oil free 1 HP Compressor with condenser, micro filter, drier mounted on tank along with high quality air tubing connection.	Brass fittings, supplied without air drier and desiccator. Inner wall of tank is not powder coated to prevent corrosion

B. M/s Kerala Surgical Equipment Co.

Rejected model:Crown-1/Crown Dent

Reason for rejection;

Clause No	Technical Specification	Compliance of the model
2.	Two Airotor points with tubing & metal nut fittings with Airotor standard head hand piece 1 No.(NSK or equivalent)	Inferior quality chuck type hand piece with standard head
7.	Sensor operated LED Overhead Light	Inferior quality with flimsy fittings. Without light intensity adjustability
8.	High quality water booster	Made of powder coated metal
13.	LED light curing unit.	Not Inbuilt
14	Oil free 1 HP Compressor with condenser, micro filter, drier mounted on tank along with high quality air tubing connection.	Brass fittings, supplied without air drier and desiccator. Inner wall of tank is not powder coated to prevent corrosion

III. Due to administrative reason the tender for **KMSCL/EP/T511/789/2023 Fully Loaded Dental Chair Electrically Operated** is cancelled.

**Dr. Shibulal A**  
General Manager