



**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [eptenders.kmscl@kerala.gov.in](mailto:eptenders.kmscl@kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

No: KMSCL/EP/1718/2024

Dated: 22.08.2024

**NOTICE**

**Sub:-** Tender No.T525/ 2024 - Results of demonstraion/ document evaluation & Date of price bid opening – Ventilator Portable - reg.

- Ref:-**
1. Tender Inviting Notice published dated 01.02.2024
  2. Technical bid opening held on 26.02.2024 @ 11.00 A M
  3. Eligible list published on 19.06.2024
  4. Reports of demonstration/ document evaluation of Ventilator Portable received on 23.07.2024

I. The reports of technical document evaluation/ demonstration of Ventilator Portable as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

SI No	Equipment	Name of the Selected Bidder	Selected Models
1.	Ventilator Portable	M/s Air Liquide Medical Systems Pvt Ltd	MONNAL T60/Air Liquide
		M/s Sree Gokulam Healthcare Pvt. Ltd.	SV 300/ Shenzhen Mindray

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below.

A. M/s Instromedix India Pvt.Ltd

Rejected model: T6/ Ambulance (Shenzhen)

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
	Comments	Not user friendly

B. M/s Parkins Enterprises

Rejected model; RV-200/ Resvent Medical

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
IV .1	220/240 V Ac 50 Hz supply. Internal battery (Li Ion) with (5-8) hours minimum operating time (hot swappable allowed)	Operating time 3 hours to 6 hours only

C. M/s Genesis Health Care

Rejected model: FT 60T/ Flight Innovations

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
9	Should be able to adjust FIO2 on the Ventilator	Can't adjust Fio2
	Comments	No facilities for mounting on trolley.

D. M/s Allied Medical Ltd.

Rejected model; MEDITECH 1700/Allied Medical

Reason for rejection.

Clause No	Technical Specification	Deviations Noted
	Comments	Not user friendly

III. The online price bid of the eligible tenderer mentioned above will be opened on 24/08.2024 @ 11.00 am at the e-tender portal.

**Dr. Shibulal A**  
General Manager