



EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/10408/2024

Dated: 18.11.2024

NOTICE

Sub:- Tender No.T544(R)/ 2024 - Results of demonstraion/ document evaluation & Date of price bid opening – various equipment - reg.

- Ref:-**
1. Tender Inviting Notice published dated 11.06.2024
 2. Technical bid opening held on 15.07.2024@ 11.00 A M
 3. Eligible list for equipment demonstration published on 21.10.2024
 4. Reports of demonstration/ document evaluation of Refrigerated Blood Transport Box received on 13.11.2024'
 5. Reports of demonstration/ document evaluation of Trial Set received on 14.11.2024

I. The reports of technical document evaluation/ demonstration of various equipment per reference no. 4th & 5th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

SI No	Equipment	Name of the Selected Bidder	Model/OEM
1.	Refrigerated Blood Transport Box	M/s Biolinx India	Biocoolbt55/Biolinx India
2.	Trial Set	M/s Hansraj Nayyar Mediquip Inc.,	H-147/ Hansraj
		M/s UNIVERSAL AGENCIES	UVTS-1004RB/ UNITECH VISION

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

1. KMSCL/EP/T544(R)/1605/2024: Refrigerated Blood Transport Box

A. M/s Tropicool car gadgets Pvt ltd

Rejected model: Tff-D60with Bb/ Tropicool Car Gadgets

Reason for rejection;

Claus no	Technical Specification	Deviations noted
4	Unit should have dual compartment with double external door	Not meeting the specification
9	Unit should have wheels and long pull rods	Not meeting the specification
11	Temperature Accuracy $\pm 1^{\circ}\text{C}$	Not meeting the specification
13	Foaming layer Thickness minimum 45mm	Not meeting the specification

2. KMSCL/EP/T544(R)/127/2024: Trial Set

A. M/s ASIANVISION

Rejected model: TRIAL LENS SET/ ASIAN VISION

Reason for rejection:

Clause No	Technical specification	Deviations noted
4.d	to 14.0 in 1.0 steps	upto 10.0 1.0 steps
4.e	16.0 to 20.0 in 2.0 steps	12.0 to 20.0 in 2.0 steps
4.g	4.0 to 6.0 in 0.5 steps	4.0 to 6.0 in 1.0 steps

III. The online price bid of the eligible tenderer mentioned above will be opened on 11.11.2024 @ 11.00 am at the e-tender portal.

Dr. Shibulal A
General Manager