



Equipment Division

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

No: KMSCL/EP/5944/2021

Dated: 22.07.2021

**NOTICE**

Sub:- Tender No. T399/ 2021 - Results of technical document evaluation/ demonstration & Date of Price Bid Opening – Mobile 100MA X-ray machine - reg.

- Ref:-
1. Tender Inviting Notice published dated 20.04.2021
  2. Technical bid opening on 18.05.2021 & 04.06.2021
  3. Demonstration/ Document evaluation of Mobile 100MA X-ray machine conducted on 16..07.2021

I. The reports of technical document evaluation/ demonstration of Mobile 100MA X-ray machine as per reference no. 3rd cited above submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidders	Selected Model/ OEM
1.	Mobile 100MA X-ray machine	M/s. Allengers Medical Systems Ltd	MARS – 4.2/ Allengers
		M/s. ADONIS MEDICAL SYSTEMS PRIVATE LIMITED	AE 100HF (Mobile)/ ADONIS

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below

A. M/s. Kerala Surgical Equipment Co

Rejected model; ALERIO SMART 4000/ IATOME

Reasons for rejection:

Clause No	Technical specification	Deviations noted
IV. b	mAs: up to 250mAs or more.	200 mAs only

III. The online price bid of the eligible tenderer mentioned above will be opened on 26.07.2021 @ 11.00 am at the e-tender portal.

Dr. Dileep Kumar S R  
 General Manager