



SUCTOR

### KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking) Thycaud.P.O., Thiruvananthapuram-14. Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600 E-mail id : <u>eptenders.kmscl@kerala.gov.in</u>

CIN: U24233KL2007SGC021616 GSTIN: 32AADCK4029M1ZK

**EP Division** 

No: KMSCL/EP/21342/2023

www.kmscl.kerala.gov.in

Dated: 28.06.2024

### NOTICE

Sub:- Tender No.T521(R)/ 2023 - Results of demonstration/ technical evaluation & Date of Price Bid Opening – various equipment - reg.

Ref:- 1. Tender Inviting Notice published dated 03.11.2023

- 2. Technical bid opening held on 27.11.2023 & 23.12.2023 @ 11.00 A M
- 3. Eligible List published on 25.01.2024
- 4. Demonstration/ document evaluation of various equipment conducted on received on 07.06.2024

I. The reports of technical document evaluation/ demonstration of various equipment as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

SI No	Equipment Name	Selected Bidders	Model/OEM	
1.	Dry Bath	M/s Aastmedi Tech Pvt Ltd	DDI 24/AAST Meditech	
		M/s Universal Agencies	RDI-SPL/ B&C	
			INSTRUMENTS	

# 1. KMSCL/EP/T521(R)/583/2023: Micro Centrifuge

A. M/s. The Engineering Science Apparatus Workshop

Rejected model: MC-50/ESAW

Reason for rejection: Sample not produced

### B. M/s.UNIVERSAL AGENCIES

Rejected model;T-16 BL, Centrifuge with Angle Rotor (24x1.5ml/2ml), max speed 13000 rpm and rcf is 16000xg/LABY INSTRUMENTS

Reason for rejection: Sample not produced

## C. M/s.LabX

Rejected model; Neya 16/REMI ELEKTROTECHNIK

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
		The equipment demonstrated is a large
		capacity multipurpose centrifuge. The
	9 20	acceleration and deceleration is too slow.
	Comments	Additional functions normally available in
Til.		micro centrifuge are not available. The
	2 2	equipment does not fully serve the purpose
	2	of micro centrifuge.

III. Due to technical reason the tender for KMSCL/EP/T521(R)/583/2023: Micro Centrifuge is cancelled.

portal.

> Dr. Shibulal A General Manager