



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/10408/2024

Dated: 07.10.2024

NOTICE

Sub:- Tender No.T544(R)/ 2024 - Results of demonstraion/ document evaluation & Date of price bid opening – various equipment - reg.

- Ref:-**
1. Tender Inviting Notice published dated 11.06.2024
 2. Technical bid opening held on 15.07.2024@ 11.00 A M
 3. Demonstration/ document evaluation of Warmer Baby conducted on 11.09.2024 & 13.09.2024
 4. Reports of demonstration/ document evaluation of Posterior Segment Surgical Microscope received on 07.10.2024

I. The reports of technical document evaluation/ demonstration of various equipment per reference no. 3rd & 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

SI No	Equipment	Name of the Selected Bidder	Model/OEM
1.	Warmer Baby	M/s Instromedix India Pvt. Ltd	Innova/M/S Bird Meditech
		M/s Phoenix Medical Systems Pvt Ltd	NWS 101 with attached bed / Phoenix Medical Systems
		M/s C CUBE ADVANCED TECHNOLOGIES.	NEONICE 3D/C CUBE ADVANCED
2.	Posterior Segment Surgical Microscope	M/s serumbiologicals	Lumera I/ Carl Zeiss

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

1. KMSCL/EP/T544(R)/141/2024: Warmer Baby

A. M/s. Dynamic International

Rejected model: DYNAMIC-S/ DYNAMIC INTERNATIONAL

Reason for rejection: Sample not produced.

B. M/s UNIVERSAL AGENCIES

Rejected model: HEAL-RWAT/HEALKEREN MEDTRONICS

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
9	The heater should automatically cut off at 38 degree Celsius irrespective of the set parameters.	Not demonstrated
18	Should be supported with easily removable side flaps.	Not Removable
19	The unit should be made of mild steel tubular structure pretreated and powder coated.	Not pretreated and powder coated.

C. M/s Alan Medical & Laboratory Products

Rejected model: P001A/MEDI WAVES INC

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
6	The probes should be detachable type and should be supplied as 2nos for each machines. The probes should be covered under warranty and CAMC.	Probe sensor very thick
16	Should be supported with easily detachable side flaps with safety lock (hook and lock mechanism)	hook and lock not available

D. M/s. Kerala Surgical Equipment Co

Rejected model; SBW 2103/Pragathy Medical Systems

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
6	The probes should be detachable type and should be supplied as 2nos for each machines. The probes should be covered under warranty and CAMC.	Probe sensor very thick
9	The heater should automatically cut off at 38 degree Celsius irrespective of the set parameters.	Not demonstrated
18	Should be supported with easily removable side flaps.	Not Removable

III. The online price bid of the eligible tenderer mentioned above will be opened on 11.10.2024 @ 11.00 am at the e-tender portal.

Dr. Shibulal A
General Manager