



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/14384/2023

Dated: 13.03.2024

NOTICE

Sub:- Tender No.T503/ 2023 - Results of demonstration/ technical evaluation & Date of Price Bid Opening – various equipment - reg.

Ref:- 1. Tender Inviting Notice published dated 09.08.2023

2. Technical bid opening held on 23.09.2023 @ 11.00 A M

3. Eligible list published on 02.11.2023

4. Demonstration/ document evaluation of various equipment is conducted on

I. The reports of technical document evaluation/ demonstration of various equipment as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models
1.	Phototherapy Unit Single Surface LED Type	M/s C Cube Advanced Technologies	NeoliteUltima Top/ C CUBE ADVANCED
		M/s Kerala Surgical Equipment Co.	Phoenix Brilliant Pro/ Phoenix Medical
		M/s Nice Neotech Medical Systems Pvt. Ltd	nice 4000 SPOT/Nice Neotech
2.	Phototherapy Unit Double Surface LED Type	M/s C Cube Advanced Technologies	NeoliteUltima/C Cube
		M/s Kerala Surgical Equipment Co	Brilliance Pro with Brilliance Bottom/Phoenix Medical
		M/s Nice Neotech Medical Systems Pvt. Ltd	nice 4000 SPOT + nice 4005/ Nice Neotech
		M/s BIOMEDIX SOLUTIONS	REX 10 BD/ IBIS MEDICAL

II. The other tenderers participated in the tender are rejected due to the reason mentioned against their name below;

1. KMSCL/EP/T503/341A/2023: Phototherapy Unit Single Surface LED Type

A. M/s Universal Agencies

Rejected model: HEAL-STP LED Stand Type LED Phototherapy unit 20nos Blue LEDs + 2 Nos White LEDs with Dig/ HEALKEREN

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
4	Irradiance level should be 35 microwatts/cm2 at 35 cms or better distance.	Not meeting the specification

B. M/s Alan Medical & Laboratory Products
 Rejected model: P003/ MEDI WAVES INC
 Reason for rejection:

Clause No	Technical Specification	Deviations Noted
4	Irradiance level should be 35 microwatts/cm2 at 35 cms or better distance.	Not meeting the specification

C. M/s R S Surgical services
 Rejected model: STP-100/RS Surgical
 Reason for rejection: Sample not produced

2. KMSCL/EP/T503/341B/2023: Phototherapy Unit Double Surface LED Type

A. M/s Universal Agencies
 Rejected model: HEAL-DSP LED Double Sided LED Phototherapy unit 40nos Blue LEDs + 2 Nos White LEDs with Digital Timer/ HEALKEREN
 Reason for rejection:

Clause No	Technical Specification	Deviations Noted
4	Irradiance level should be 35 microwatts/cm2 at 35 cms or better distance.	Not meeting the specification

B. M/s Alan Medical & Laboratory Products
 Rejected model: P003/ MEDI WAVES INC
 Reason for rejection:

Clause No	Technical Specification	Deviations Noted
4	Irradiance level should be 35 microwatts/cm2 at 35 cms or better distance.	Not meeting the specification

C. M/s R S Surgical services
 Rejected model; STP-100/RS Surgical
 Reason for rejection: Sample not produced

III. The online price bid of the eligible tenderer mentioned above will be opened on 15.03.2024 @ 11.00 am at the e-tender portal.