



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)
Thycaud.P.O., Thiruvananthapuram-14.
Tele Fax No. 0471-2945647
0471-2945646, 0471-2945600

E-mail id: eptenders.kmscl@kerala.gov.in CIN: U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

Dated: 18.12.2023

No: KMSCL/EP/14386/2023

NOTICE

Sub:- Tender No.T505(R)/ 2023 - Results of demonstration/ document evaluation & Date of price bid opening – Battery Operated Power Drill with Reamer - reg.

Ref:- 1. Tender Inviting Notice published dated 09.08.2023

- 2. Technical bid opening held on 04.09.2023 @ 11.00 A M
- 3. Eligible list published on 20.10.2023
- 4. Reports of demonstration/ document evaluation of Battery Operated Power Drill with Reamer received on 14.12.2023

I. The reports of technical document evaluation/ demonstration of Battery Operated Power Drill with Reamer as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

SI. No	ltem	Name of the Selected Bidder	Model/OEM
1	Battery Operated Power Drill with	M/s Bions Medical Systems	Battery Power Line II/ Johnson &
	Reamer	Pvt Ltd	Johnson

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

A. M/s Neoptic Surgicals

Rejected model: System 8 Power tool/ Stryker

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
1.b.	Single Trigger	No single trigger
6.a	For dia 1.5 to 4.0 mm Wires	Not meeting the specification
8.b	Long battery lifetime (> 1000 charging Cycles)	Not meeting the specification
11 D.3	Flexible Reamer shaft 7.0 mm dia. Reaming depth 470mm, made of Nitinol used both clockwise & counter clock wise direction. Length 470mm	The reusable detachable reamer has only one direct channel