



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/2412/2023

Dated: 29.05.2023

NOTICE

Sub:- Tender No.T489(R)/ 2023 - Results of demonstration/ technical evaluation & Date of Price Bid Opening – various equipments - reg.

- Ref:-**
1. Tender Inviting Notice published dated 23.02.2023
 2. Technical bid opening held on 13.03.2023, 18.03.2023, 24.03.2023, 27.03.2023 & 03.04.2023 @ 11.00 A M @ 11.00 A M
 3. Eligible list published on 26.04.2023
 5. Demonstration/document evaluation of Iron Cot conducted on 04.05.2023 & 05.05.2023
 6. Reports of demonstration/ document evaluation received on 25.05.2023

I. The reports of technical document evaluation/ demonstration of various equipment as per reference no. 5th & 6th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Item	Name of the Selected Bidder
1.	Iron Cot	M/s.Kerala Surgical Equipment Co.
		M/s.Midmark (India) Private Limited
		M/s.STEEL INDUSTRIALS KERALA LIMITED
		M/s.Trivandrum Surgicals and Lab Supplies
2	Video Thoracoscope with HD Camera & Accessories	M/s Neoptic Surgicals

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below:

A. KMSCL/EP/T489(R)/364/2023: Iron Cot

1. M/s. HOSPITAL BEDS
Reason for rejection: Sample not produced
2. M/s.Ashly Furnishing Co Pvt Ltd
Reason for rejection: Sample not produced
3. M/s.Alan Medical & Laboratory Products
Reason for rejection :

Clause No	Technical Specification	Deviations Noted
4	The top sheet shall be 18 gauge(1.2mm) thickness CRCA sheet double bend of height minimum reinforced with channels of size minimum 75mm x 30mm flange welded inside the lengthwise frame bends.	Not meeting the specification

