



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/4151/2024

Dated: 06.11.2024

NOTICE

Sub:- Tender No.T535/ 2024 - Results of demonstration/ technical evaluation & Date of Price Bid Opening – various equipment - reg.

- Ref:-**
1. Tender Inviting Notice published dated 05.03.2024
 2. Technical bid opening held on 20.05.2024, & 01.06.2024 @ 11.00 A M
 3. Eligible list published on 28.09.2024
 4. Demonstration/ document evaluation of various equipment conducted on 22.10.2024

I. The reports of technical document evaluation/ demonstration of various equipment as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Model
1.	Automated Immune Assay Analyzer Model A	M/s.CPC Diagnostics Pvt Ltd	IFlash 1200/Shenzhen Yhlo Biotech
		M/s.Siva Management Services	CL 900i Automated Chemiluminescence Immunoassay Analyzer/Shenzhen Mindray
		M/s.TRICORE MEDICAL SYSTEMS	Mispa-i200/Agappe
		M/s.ABC DIAGNOSTICS PVT. LTD	VIDAS KUBE/BIOMERIEUX
2.	Fully Automated Immune Assay Analyzer Model D	M/s. Siva Management Services	CL 1200i Fully Automated Chemiluminescence Immunoassay Analyzer/Shenzhen Midray
		M/s. TRICORE MEDICAL SYSTEMS	Mispa-i180/ Maccura

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

A. KMSCL/EP/T535/436D/2024: Fully Automated Immune Assay Analyzer Model D

1. M/s. Beckman Coulter India Private Limited
Rejected model: Access 2/Beckman Coulter
Reason for rejection:

Clause No	Technical Specification	Deviations Noted
12	The reagent on- board and calibration stability of reagent should be a minimum of 14 days for Thyroid parameters and 28 or more for all other parameters	Not meeting the specification

2. M/s. Genworks Health Private Limited
Rejected model: Maglumi X3/Shenzhen
Reason for rejection: Sample not produced

3. M/s. Instromedix India Pvt. Ltd

Rejected model: Magicl 6000i/Getein Biotech

Reason for rejection: Sample not produced

III. The online price bid of the eligible tenderer mentioned above will be opened on 08.11.2024 @ 11.00 am at the e-tender portal.

Dr. Shibulal A
General Manager