

**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : etenders_ep@kmscl.kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

Equipment Division**www.kmscl.kerala.gov.in****No: KMSCL/EP/1029/2019****Dated: 23.12.2020****NOTICE**

Sub:- Tender No. T351(R)/2019- Results of technical document evaluation/ demonstration & Date of Price Bid Opening – various equipments - reg.

- Ref:-**
1. Tender Inviting Notice published dated 22.11.2019
 2. Technical bid opening on 20.12.2019 & 28.12.2019, 31.12.2019 @ 11.00 A M.
 3. Clarification notice published on 28.12.2019.
 4. Eligible list published on 30.01.2020
 5. Reports of Demonstration/Document evaluation of Video Intubating Bronchoscope Model B submitted by the technical committee
 6. Reports of Demonstration/Document evaluation of Hand Held Slit Lamp submitted by the technical committee
 7. Reports of Demonstration/Document evaluation of TMJ Arthroscopy Set submitted by the technical committee

I. The reports of technical document evaluation/ demonstration of various items as per reference no. 5th, 6th & 7th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

SI No	Equipment Name	Name of the selected bidder	Selected Model/OEM
1.	Video Intubating Broncho scope Model B	M/s Neoptic Surgicals	iS3-C/shenzhen Insighters
		M/s Lakshmi Surgicals	Karl Storz Video Intubating Bronchoscope ModelB/Karl Storz
2.	Hand Held Slit Lamp	M/s Suraj Hi-Tech Pvt.Ltd.,	XL-1/Shin Nippon By Rexxam,
3	TMJ Arthroscopy Set	M/s Lakshmi Surgicals	Karl Storz Tmj Arthroscopy Set/Karl Storz

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

1. KMSCL/EP/T351(R)/110F/2019: Hand Held Slit Lamp

A. M/s Appasamy Associates Private Limited

Rejected Model: Portable/ Appasamy Associates

Reason for rejection:

SI No	Technical Specification	Deviations Noted
3. d	Slit rotation : 30 degrees	This feature is not properly available
		The slit rotation of the offered model was not adequate and was not properly aligned for

	Comment	viewing the anterior segment structure during the demonstration.
--	----------------	--

II.The online price bid of the eligible tenderer mentioned above will be opened on 26 .12.2020 @ 11.00 am at the etender portal.

Dr. Dileep Kumar S R
General Manager