



**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [etenders\\_ep@kmscl.kerala.gov.in](mailto:etenders_ep@kmscl.kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

No: KMSCL/EP/16693/2021

Dated: 22.06.2022

**NOTICE**

**Sub:-** Tender No.T422(R)/ 2021 - Results of demonstration/ technical evaluation & Date of Price Bid Opening – Automatic External Defibrillator (AED) - reg.

**Ref:-** 1. Tender Inviting Notice published dated 01.11.2021

2. Technical bid opening held on 22.11.2021 @ 11.00 A M

3. Eligible list for demonstration published on 11.03.2022

4 Demonstration/ document evaluation of Automatic External Defibrillator (AED) conducted on 30.05.2022

I. The reports of technical document evaluation/ demonstration of Automatic External Defibrillator (AED) as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Item	Name of the Selected Bidder	Selected Model
1.	Automatic External Defibrillator (AED)	M/s. ANAMDEV ENGINEERS	Heartstart FRX/Philips
		M/s. BIOCARE DIAGNOSTICS	AED 3100/NIHON KOHDEN

II. The other tenderers participated in the tender are rejected due to the reason mentioned against their name below;

A. M/s.UNIVERSAL AGENCIES

Rejected Model:: HR-501/RADIANQBIO

Reason for rejection:

Clause no	technical Specification	Deviations noted
13	Should have 1 m drop test passed, along with at least IP 55 rating.	IP 54 rating

III. The online price bid of the eligible tenderer mentioned above will be opened on 23.06.2022 @ 11.00 am at the e-tender portal.

**Dr. S S Joy**

General Manager