

**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : etenders_ep@kmscl.kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

Equipment Divisionwww.kmscl.kerala.gov.in**No: KMSCL/EP/2136/2021****Dated: 10.08.2021****NOTICE**

Sub:- T393/2021 - T393/2021 - Results of technical document evaluation/ demonstration & Date of Price Bid Opening - Basic Dental Chair (Electrically Operated) - reg.

Ref:- 1. Tender Inviting Notice published dated 11.02.2021
2. Technical bid opening on 15.03.2021 ,22-03-2021@ 11.00 A M
3. Eligible list for equipment demonstration published on 16.04.2021
4. Demonstration/ document evaluation of Basic Dental Chair (Electrically Operated) conducted on 29.07.2021 & 10.08.2021

I. The reports of technical document evaluation/ demonstration of various equipments as per reference no.4th cited above were submitted by the technical committee appointed for the purpose. The result notice of Instrument Trolley was published earlier but due to oversight the eligible list published is incomplete. The revised details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models
1.	Basic Dental Chair (Electrically Operated)	M/s. CONFIDENT DENTAL EQUIPMENTS LTD	Wave Dental/ Confident
		M/s. DENTLINES	Electrica/ Clinix Inteligent
		M/s Trichur Surgicals	TS 431/ Trichur Surgicals

II. The other tenderers participated in the tenderer are rejected due to the reasons mentioned against thier name below:

A. M/s. Kerala Surgical Equipment Co

Rejected Model: Medi Dent1/ Medik Enterprises

Reason for rejection:

Clause No	Technical Specification	Deviations noted
2	Auto water connection for spittoon	Design of the water connection to the spittoon is flimsy. Area of spittoon demonstrated is too small and it may contribute spillage and contamination of the surrounding area.
	Comments	The hand rest of the chair is not positioned properly for comfortable patient seating. The design of the chair demonstrated is not ergonomic. Overall workman ship of the chair demonstrated is poor.

III.The online price bid of the eligible tenderers mentioned above will be opened on 12.08.2021 @ 11.00 am at the etender portal.

Dr. Dileep Kumar S R
General Manager