



EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/4229/2024

Dated: 05.11.2024

NOTICE

Sub:- Tender No.T536/ 2024 - Results of demonstration/ document evaluation & Date of pricebid opening – Cell Saver - reg.

- Ref:-**
1. Tender Inviting Notice published dated 05.03.2024
 2. Technical bid opening held on 06.05.2024 @ 11.00 A M
 3. Eligible list for demonstration published dated 24.08.2024
 4. Reports of demonstation/ document evaluation of Cell Saver received on 01.11.2024

I. The reports of technical document evaluation/ demonstration of Cell Saver as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

SI. No	Equipment	Name of the Selected Bidder	Selected Models
1.	Cell Saver	M/s Medibright Surgicals	XTRA/ Livanova Deutschland

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

A. M/s Haemolab Equipments

Rejected model; CATSmart/ Fresenius Kabi

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
6	Washing bowls should be in minimum 4 different sizes for various kinds of patients- 75ml, 125ml, 175ml, 225ml.	Not meeting the specification
7	Should have an option of doing Blood sequestration process for PPP, PRP1, and PRP2 using same bowl set.	Not meeting the specification
9	It should have inbuilt memory for storing up to 10,000 cases. Should have inbuilt printer with permanent printouts.	Not meeting the specification
17	Should have a facility for sequestration of whole blood into Platelets & Plasma using the same disposable Bowl set as for Intra-operative blood processing	Not meeting the specification

III. The online price bid of the eligible tenderer mentioned above will be opened on 07.11.2024 @ 11.AM at the etender portal.

