



**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [eptenders.kmscl@kerala.gov.in](mailto:eptenders.kmscl@kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

No: KMSCL/EP/14384/2023

Dated: 14.03.2024

**NOTICE**

**Sub:-** Tender No.T503/ 2023 - Cancellation Notice – Warmer Baby - reg.

**Ref:-** 1. Tender Inviting Notice published dated 09.08.2023

2. Technical bid opening held on 23.09.2023 @ 11.00 A M

3. Eligible list published on 02.11.2023

4. Demonstration/ document evaluation of Warmer Baby is conducted on 05.02.2024 & 16.02.2024

I. The reports of technical document evaluation/ demonstration of Warmer Baby as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models
1.	Warmer Baby	M/s Kerala Surgical Equipment Co.	NWS 101/ Phoenix Medical

II. The other tenderers participated in the tender are rejected due to the reason mentioned against their name below;

A. M/s Universal Agencies

Rejected model: HEAL-RWAT/ HEALKEREN MEDTRONICS

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
	Comments	baby mattress is thin and monitor stand is not sturdy

2. M/s Alan Medical & Laboratory Products

Rejected model; POO1A/MEDI WAVES INC

Reason for rejection; Sample not produced

3. M/s Nice Neotech Medical Systems Pvt. Ltd

Rejected model : nice 2007SG/ Nice Neotech

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
6	The probes should be detachable type and should be supplied as 2nos for each machines. The probes should be covered under warranty and CAMC.	Probes not covered under warranty

III. Due to administrative reason the tender for **KMSCL/EP/T503/141/2023: Warmer Baby** is cancelled.

**Dr. Shibulal A**  
General Manager