



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/21342/2023

Dated: 25.06.2024

NOTICE

Sub:- Tender No.T521(R)/ 2023 - Results of demonstration/ technical evaluation & Date of Price Bid Opening – various equipment - reg.

- Ref:-**
1. Tender Inviting Notice published dated 03.11.2023
 2. Technical bid opening held on 27.11.2023 & 23.12.2023 @ 11.00 A M
 3. Eligible List published on 25.01.2024
 4. Demonstration/ document evaluation of various equipment conducted on received on 04.06.2024

I. The reports of technical document evaluation/ demonstration of various equipment as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

SI No	Equipment Name	Selected Bidders	Model/OEM
1.	Micro-Ear Surgery Instruments	M/s Trivandrum Surgicals and Lab Supplies	CHATTERJEE/ CHATTERJEE

1. KMSCL/EP/T521(R)/990/2023: Septoplasty Set

A. M/s. Trivandrum Surgicals and Lab Supplies

Rejected model: CHATTERJEE/ CHATTERJEE

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
	Comments	Septoplasty instruments specified in the tender is an incomplete set. Fess instruments included in this tender do not come under septoplasty set. Hence the committee decides to retender with revised specification.

2. M/s Manchel Marketing Company

Rejected model; KALELKER Instruments, H LED /Kalelker

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
4	Mahadevaiya's self retaining endomeatal retractor	Not meeting the specification
5	Farabeuf periosteal elevator, SS - Tip 11 mm wide	Not demonstrated
6	Lempert's periosteal elevator (3 mm) -18 cms long	Not demonstrated
7	Cell seeker 45° with ball end - 15.5 cms long	Not meeting the specification

12	more cup (scope up thing, star cup 100 mm, 8cms working length	Not meeting the specification
13	Myringotomy knife adjustable (160 mm)	Not demonstrated
16	Rosen round knife, 45 degree angled, cutting - 1 mm dia. - 16 cms long	Not demonstrated
24.b	Fisch Perforator 0.2mm tip	Not demonstrated
24.c	Fisch Perforator 0.4mm tip	Not demonstrated
24.d	Fisch Perforator 0.6mm tip	Not demonstrated
25.b	Fisch Perforator 0.2mm tip	Not demonstrated
25.c	Fisch Perforator 0.4mm tip	Not demonstrated

2. M/s Radha Meditech

Rejected model; Santosh Surgicals

Reason for rejection: Sample not produced

III. Due to technical reason the tender for **KMSCL/EP/T521(R)/990/2023: Septoplasty Set** is cancelled.

IV. The online price bid of the eligible tenderer mentioned above will be opened on 2806.2024 @ 11.00 am at the e-tender portal.

Dr. Shibulal A
General Manager