



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/662/2024

Dated: 24.07.2024

NOTICE

Sub:- Tender No.T522/ 2024 - Results of demonstration/ technical evaluation & Date of Price Bid Opening – various equipment - reg.

- Ref:-**
1. Tender Inviting Notice published dated 16.01.2024
 2. Technical bid opening held on 31.01.2024, 05.02.2024 @ 11.00 A M
 3. Eligible list published on 19.06.2024
 4. Demonstration/ document evaluation of various equipment conducted on 18.07.2024

I. The reports of technical document evaluation/ demonstration of various equipment as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl No	Equipment	Name of the Selected Bidder	Selected Models
1.	Ambu Bag – Infant	M/s Trivandrum Surgicals and Lab Supplies	AMBU BAG/FLEXICARE
		M/s Biomedical Engineering Company	Oval Plus Silicon/Oval Plus Silicon
2.	Ambu Bag – Child	M/s Trivandrum Surgicals and Lab Supplies	AMBU BAG/FLEXICARE
		M/s Biomedical Engineering Company	Oval Plus Silicon/Oval Plus Silicon
3.	Ambu Bag – Adult	M/s Trivandrum Surgicals and Lab Supplies	AMBU BAG/FLEXICARE
		M/s Biomedical Engineering Company	Oval Plus Silicon/Oval Plus Silicon
4.	Nebulizer ultrasonic	M/s Kerala Surgical Equipment Co.	NEU-780/Omron Health Care Co.Ltd
5.	Sphygmomanometer Model	Stand M/s Trivandrum Surgicals and Lab Supplies	DIAMOND BPDG 534/INDUSTRIAL ELECTRONIC &ALLIEDPRODUCTS

II. The other tenderers participated in the tender are rejected due to the reason mentioned against their name below;

A. KMSCL/EP/T522/1/2024 : Ambu Bag – Infant

1. M/s Alan Medical & Laboratory Products

Rejected model: AMLP BAGS/ALAN

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
5	Should have an expiratory resistance of 2.2cms of water	Not demonstrated
6	Should have an inspiratory resistance of 3.3cms of water.	Not demonstrated

2. M/s UNIVERSAL AGENCIES

Rejected model; SKU 40002/NDOSURGICALS

Reason for rejection: Sample not produced

3. M/s Lakshmi Agencies

Rejected model: Medi Safe International

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
5	Should have an expiratory resistance of 2.2cms of water	Not demonstrated
6	Should have an inspiratory resistance of 3.3cms of water.	Not demonstrated

4. M/s Kerala Surgical Equipment Co

Rejected model; Continental

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
	Comments	Manufacture details not provided

B. KMSCL/EP/T522/1/2024 Ambu Bag – Child

1. M/s Alan Medical & Laboratory Products

Rejected model: AMLP BAGS/ALAN

Reason for rejection;

Clause No	Technical Specification	Deviations Noted
5	Should have an expiratory resistance of 2.2cms of water	Not demonstrated
6	Should have an inspiratory resistance of 3.3cms of water.	Not demonstrated
7	It should have controlled flow rates and ventilation, and with reduced airway pressure.	Not demonstrated

2. M/s UNIVERSAL AGENCIES

Rejected model; SKU 40001/INDOSURGICALS

Reason for rejection; Sample not produced

3. M/s Lakshmi Agencies

Rejected model; Medi Safe International

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
5	Should have an expiratory resistance of 2.2cms of water	Not demonstrated
6	Should have an inspiratory resistance of 3.3cms of water.	Not demonstrated

4. M/s Kerala Surgical Equipment Co

Rejected model: Continental

Reason for rejection;

Clause No	Technical Specification	Deviations Noted
	Comments	Manufacture details not provided

C. KMSCL/EP/T522/1/2024 Ambu Bag – Adult

1. M/s Alan Medical & Laboratory Products

Rejected model: AMLP BAGS/ALAN

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
5	Should have an expiratory resistance of 2.2cms of water	Not demonstrated
6	Should have an inspiratory resistance of 3.3cms of water.	Not demonstrated
7	It should have controlled flow rates and ventilation, and with reduced airway pressure.	Not demonstrated

2. M/s UNIVERSAL AGENCIES

Rejected model; SKU 4000/INDOSURGICALS

Reason for rejection; Sample not produced

3. M/s Lakshmi Agencies

Rejected model; Medi Safe International

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
5	Should have an expiratory resistance of 2.2cms of water	Not demonstrated
6	Should have an inspiratory resistance of 3.3cms of water.	Not demonstrated

4. M/s Kerala Surgical Equipment Co

Rejected model; Continental

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
	Comments	Manufacture details not provided

D. KMSCL/EP/T522/253/2024; Nebulizer ultrasonic

1. M/s Lakshmi Agencies

Rejected model: NEBTIME UN600A (2.5MHZ)/ELMASLAR MEDIKAL

Reason for rejection;

Clause No	Technical Specification	Deviations Noted
6	Medication cup capacity should have capacity of maximum 8ml.	Not demonstrated
12	Should have safety certificate from a competent authority CE / FDA (US) or any other equivalent standards.	No valid certificate

E. KMSCL/EP/T522/113/2024; Sphygmomanometer Stand Model

1. M/s Kerala Surgical Equipment Co

Rejected model: Lifeline

Reason for rejection: Sample not produced

2. M/s Lakshmi Agencies

Rejected model: kifzi

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
11	The tubes should be fitted with male and female leur connectors.	Not leur connectors
12	The inflating bulb should be soft and should not have any joints or ridges.	inflating bulb was having ridges

III. The online price bid of the eligible tenderer mentioned above will be opened on 29.07.2024 @ 11.00 am at the e-tender portal.

Dr. Shibulal A
General Manager