



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/20599/2023

Dated: 20.06.2024

NOTICE

Sub:- Tender No.T518/ 2023 - Cancellation Notice – Hand held non mydriatic camera - reg.

Ref:- 1. Tender Inviting Notice published dated 28.11.2023

2. Technical bid opening held on 15.12.2023 @ 11.00 A M

3. Eligible list published on 07.03.2024

4. Results of demonstration/ document evaluation of Hand held non mydriatic camera published on 31.05.2024

5. Complaint mail received on 03.06.2024

6. Report of technical committee received on 18.06.2024

I. With reference to the demonstration result notice published for Hand held non mydriatic camera a complaint was received against the model FC 161/ SHANGHAI offered by M/s Lumen Mediworks Pvt. Ltd. The complaint received was verified by the technical committee and the findings of the committee are as follows;

Tender Spec Sno	Tender specification	Remarks of FC161 offered.	Remarks of the technical committee
10	Levels of Illumination: ISO15004 safety assured illumination levels controllable through software	Does not meet the specification	The representation submitted by the bidder was verified in detail. It is observed that both the designs are different. Since the overall purpose is met, the committee unanimously selected the offer of M/S. Lumen Mediworks Private Limited by accepting the minor deviation, as the deviation is due to the difference in design technology. As M/s Remidio Innovative Solutions Private Limited represented against the selection, it is recommended to retender by changing the specification by accommodating both technologies.
11	Regulator certificate: CE, IEC6060-1, IEC6060-2, IEC 62304, IEC62133, ISO15004-1-2,ISO10940,USFDA510K	Does not meet the specification	
12	AC Power adaptor:9V,2A external CE marked medical grade adaptor.	Does not meet the specification	
13	Camera & Display Interface: Smartphone Camera with integrated smartphone-based touch display	Does not meet the specification	
14	Camera resolution: Equal or greater than 80 lp/mm at imaging plane traceable to ISO10940 with at least 8 MP Pixel resolution.	Does not meet ISO 10940 condition.	
15	Form & Construction: Fixed smartphone holder fixed to optical interface.	Does not meet the specification.	
16	Weight & Dimension (WxDxH) 1.5kg & 68 X 248 X 234mm	Does not meet the specification	
17	Battery & back up: 7.4V, at least 1400 mAh Li-ion rechargeable battery.	Does not meet the specification.	
19	Acquisition Software: iOS Operating System based, with built in patient management Software covering Patient Name, ID, Age, DOB, Hospital Name, Doctor Name, Address fields, ISO, focus Zoom, LED intensity controls, manual focusing on LIVE view, color, and IR	Does not have iOS operating system.	

	photo acquisition mode.	
21	Montage Software: Automated montage software on the smart phone.	Does not have smart phone in the device.
22	Connectivity options: Bluetooth, Wi-Fi, SIM Slot with pre-loaded smartphone apps using these connectivity features, HIPAA compliant cloud sync.	Does not meet the specification
23	Illumination technology: Annular Illumination technology with black absorber to eliminate corneal reflexes.	Does not meet the specification.
24	Device clinical validation: Technology validated in 7 segment ETDRS study.	Does not meet the specification

II. Due to administrative reason the tender for **KMSCL/EP/T518/994/2023: Hand held non mydriatic camera** is cancelled.

Dr. Shibulal A
General Manager