



EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/2412/2023

Dated: 11.09.2023

NOTICE

Sub:- Tender No.T489(R)/ 2023 - Results of demonstration/ technical evaluation & Date of Price Bid Opening – various equipment - reg.

Ref:- 1. Tender Inviting Notice published dated 23.02.2023

2. Technical bid opening held on 13.03.2023, 18.03.2023, 24.03.2023, 27.03.2023 &.03.04.2023@ 11.00 A M @ 11.00 A M

3. Eligible list published on 26.04.2023

4. Demonstration/ document evaluation of various equipment conducted on .03.08.2023 & 16.08.2023

I. The reports of technical document evaluation/ demonstration of various items as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models
1.	Blood Bank Refrigerator (300 Litre)	M/s.KOCHIN INLAB EQUIPMENTS INDIA PVT LTD	BR160/TERUMO PENPOL PVT LTD
2.	Electronic Weighing Scale for Blood Bank	M/s.KOCHIN INLAB EQUIPMENTS INDIA PVT LTD	CS 300/TERUMO PENPOL PVT LTD
3	Apherisis Machine	M/s KOCHIN INLAB EQUIPMENTS INDIA PVT LTD	Spectra Optia/Terumo BCT,USA
4	Vertical Plasma Freezer - 40 Degree (Deep Freezer-40 Degree)	M/s.KOCHIN INLAB EQUIPMENTS INDIA PVT LTD	DF40U/ Terumo Penpol

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below:

A. KMSCL/EP/T489(R)/151/2023: Blood Bank Refrigerator (300 Litre)

1. M/s. Remi Elektrotechnik Ltd

Rejected model: BR120 With VS 02

Reasons for rejection:

Clause No	Technical Specification	Deviations Noted
13	Should have a vertical cabinet.	Horizontal Cabinet
20	Should have safety certificate from a competent authority CE issued by a notified body registered in European commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.	No valid certificate

2. M/s. Labline Instruments

Rejected model: BBR-120

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
11	Should have visual, audible indication for door open, high and low temperature and power on.	No visual alarm for door open
20	Should have safety certificate from a competent authority CE issued by a notified body registered in European commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.	No valid certificate

3. M/s. Biocare Diagnostics

Rejected model: GBBR – 120

Reason for rejection: Sample not produced

4. M/s. New India Surgical and Chemicals

Rejected model :BBR-160

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
16	Should have sensors for activating automatic defrost cycles to minimize the frost build up.	Automatic defrost not available
20	Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid	No Valid Certificate

II. KMSCL/EP/T489(R)/267A/2023: Electronic Weighing Scale for Blood Bank

1. M/s. Scientific Medical Equipments (Kerala) Pvt

Rejected Model: HTBBW1/HERCULES

Reason for rejection: Sample not produced

III. KMSCL/EP/T489(R)/638/2023: Apherisis Machine

1. M/s. Fresenius Kabi India Pvt Ltd

Rejected model: Comtec

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
17	Simple kit installation	No, Kit installation is not simple. 27 user interventions needed to start the procedure. Priming takes 6 minutes.
19	Yield of 3.3x10 ¹¹ even at pre-count of 2.65 lakhs per micro litre	Kit loading failed

IV. KMSCL/EP/T489(R)/263/2023: Vertical Plasma Freezer - 40 Degree (Deep Freezer-40 Degree)

1. M/s Genist Technocracy Pvt Ltd

Rejected model: GTP-003/Genist Technocracy

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
3	Number of Trays: 3 adjustable stainless steel trays	No
23	Temperature Recording: Seven days temperature recording shall be available. Temperature recording chart (300nos) and ink pen (5 nos) shall be supplied free of cost	Not Available
24	Battery backup for display and recorder shall be available for at least 2 hours.	Not Available
26	Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.	No Valid Certificate

2. M/s Remi Elektrotechnik Ltd

Rejected model; RPF 336 With VS 03/ Remi Elektrotechnik

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
10	Door Insulation: 125mm thick PUF insulated door with rubber gasket sealing	No,Black PUF attached to the door can have crevices which makes cleaning difficulty and compromise sterility of equipment.
26	Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.	No Valid Certificate

3. M/s Labline Instruments

Rejected model: VDF 2853/ Labline Instruments

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
26	Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.	No Valid Certificate

4. M/s Biocare Diagnostics

Rejected model: GDF – 320P/ GENESYS

Reason for rejection : Sample not produced

5. M/s New India Surgical and Chemicals

Rejected model: URC-V-325-2/ Cryo Scientific

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
26	Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.	No Valid Certificate

6. M/s Biolinx India

Rejected Model: TPF- 300V/ Jeshra Instruments

Reason for rejection: Sample not produced

III. The online price bid of the eligible tenderer mentioned above will be opened on 13.09.2023 @ 11.00 am at the e-tender portal.

Dr. Shibulal A
General Manager

