



**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [etenders\\_ep@kmscl.kerala.gov.in](mailto:etenders_ep@kmscl.kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

**Equipment Division**

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

**No: KMSCL/EP/11987/2023**

**Dated: 19.11.2023**

**NOTICE**

**Sub:-** Tender No.T498/ 2023 - Cancellation Notice – TMT Machine Model B - Reg:-

**Ref:-** 1. Tender Inviting Notice published dated 05.07.2023

2. Technical bid opening held on 07.08.2023 @ 11.00 A M

3. Eligible list published on 20.09.2023

4. Reports of demonstration/ technical document evaluation of TMT Machine Model B received on 08.11.2023

I. The reports of technical document evaluation/demonstration of TMT Machine Model B as per reference no.4th cited above were submitted by the Technical Committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models
1.	TMT Machine Model B	M/sBiomedical Engineering Company	Cardiosoft with T 1000 ST Treadmill/ GE HEALTHCARE

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below:

1. M/s Allengers Global Healthcare Pvt Ltd.

Rejected model:Gemini-A-DX/ Allengers Global

Reason for rejection::

Clause No	Technical Specification	Technical compliance of the model
30	The system must provide a detachable/fixed patient ECG acquisition module with replaceable patient lead wires	Lead wires are not replaceable

2. M/s Universal Agencies

Rejected model: EC 12R/S/ LABTECH Ltd ,HUNGARY

Reason for rejection: Sample not produced

III. Due to administrative reason the tender for **KMSCL/EP/T498/272C/2023 : TMT Machine Model B** is cancelled

**Dr. Shibulal A**

