



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)
Thycaud.P.O., Thiruvananthapuram-14.
Tele Fax No. 0471-2945647
0471-2945646. 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in
CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

www.kmscl.kerala.gov.in

Dated: 31.05.2024

EP Division

No: KMSCL/EP/673/2024

NOTICE

Sub:- Tender No.T523(R)/ 2023 - Results of demonstration/ technical evaluation & Date of Price Bid Opening – various equipment - reg.

Ref:- 1. Tender Inviting Notice published dated 16.01.2024

- 2. Technical bid opening held on 05.02.2023 & 09.02.2024 @ 11.00 A M
- 3. Eligible list published on 07.03.2024
- 4. Reports of demonstration/ document evaluation of AB Scan received on 29,05,2024
- 5. Demonstration/ document evaluation of Retractor Set conducted on 25,05,2024
- I. The reports of technical document evaluation/ demonstration of various equipment as per reference no. 4th & 5th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

SI. No	Equipment	Name of the Selected Bidder	Selected Models
1.	AB Scan	M/s Appasamy Associates	Marvel II AB-Scan/Appasamy
		Pvt Ltd	Associates

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below:

A. KMSCL/EP/T523(R)/1146/2023: Retractor Set

1. M/s Kerala Surgical Equipment Co

Rejected Model; Associated Surgical

Reason for rejection;

Clause No	Technical Specification	Deviations Noted
5	The retractor blade should be 6" long and 1.5" wide, handle length should be 6".	The retractor blade is 4" only
6	The anterior wall retractor should have flat blade 4"long and 1.5"wide and a handle length of 6".	Not demonstrated

III. Due to technical reason the tender for KMSCL/EP/T523(R)/1146/2023: Retractor Set is cancelled.

IV. The online price bid of the eligible tenderer mentioned above will be opened on 04.05.2024 @ 11.00 am at the e-tender portal.