



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/11988/2023

Dated: 25.01.2024

NOTICE

Sub:- Tender No.T499/ 2023 - Results of demonstration/ document evaluation & Date of price bid opening - A- Scan - reg.

- Ref:-**
1. Tender Inviting Notice published dated 05.07.2023
 2. Technical bid opening on 24.07.2023 @ 11.00 A M
 3. Reports of demonstration/ document evaluation of A- Scan received on 23.01.2024

I. The reports of technical document evaluation/ demonstration of A- Scan as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

SI No	Equipment Name	Name of the selected bidder	Selected model
1.	A- Scan	M/s UNIVERSAL AGENCIES	A-SCAN PLUS CONNECT/ KEELER LTD. (ACCUTOME)
		M/s APPASAMY ASSOCIATES PRIVATE LIMITED	Appascan Max/ Appasamy

II. The other tenderers participated in the tender are rejected due to the reason mentioned against their name below;

A. M/s AKAS TECHNOLOGIES

Rejected model: BIOSCALE/ AKAS TECHNOLOGIES

Reason for rejection:

Clause No	Technical specification	Deviations noted
1	Should have easy readable LCD display	Not meeting the specification
3	Should have a probe frequency of 10Mhz or more with internal fixation light	Not meeting the specification
5	Should have automatic, manual, PMMA, Silicone, Acrylic modes.	Not meeting the specification
6	Should provide IOL formulas such as Holladay II, Hoffer Q, SRK/T, Haigis etc and should have accepted post refractive formulas like shammas or equivalent.	Not meeting the specification

III. .The online price bid of the eligible tenderer mentioned above will be opened on 31.01.2024 @ 11.00 AM at the e-tender

Dr. Shibulal A
General Manager