



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)
Thycaud.P.O., Thiruvananthapuram-14.
Tele Fax No. 0471-2945647
0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/4144/2024

Dated: 21.10.2024

NOTICE

Sub:-Tender No.T531/ 2024 - Results of demonstration/ technical evaluation & Date of Price Bid Opening – Biphasic Defibrillator Monitor Pacer CPR Feedback - reg.

Ref:- 1. Tender Inviting Notice published dated 05.03.2024

- 2. Technical bid opening held on 20.05.2024 @ 11.00 A M
- 3. Eligible list published on 20.07.2024
- 4. Reports of demonstration/ document evaluation of Biphasic Defibrillator Monitor Pacer CPR Feedback received on 17.10.2024

I. The reports of technical document evaluation/ demonstration of Biphasic Defibrillator Monitor Pacer CPR Feedback as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

SI. No	Equipment			Name of the Selected Bidder			Selected Models		
1.	Biphasic Defibrillator Monitor						BeneHeart D60/ Shenzhen Mindary		
				M/s Aeon			Corpuls3/	24	GS
				MedicalPrivate Limited			Elektromedizinische		

II. The other tenders participated in the tender are rejected due to the reasons mentioned against their name below;

A. M/s BIOCARE DIAGNOSTICS

Rejected model: EMS 1052/ NIHON KOHDEN

Reason for rejection:

Clause No	Technical Specification	Compliance of the model			
	Having facility of integrated real				
	CPR feedback system. Ability to				
	measure chest compression rate				
	and depth in real time with visual				
	on-screen indicator and audible	ja			
	feedback prompt the user such as				
	push harder, good compression.	Not mosting the			
7	There should be facilty to retrive	Not meeting the			
	the data along with all the	specification.			
	measured parameters during the				
	session. Retrieved data should be				
	on a second by second basis.				
	Provision to calculate chest				
	compression fraction, interuptions,				
	defibrillation indication preferred.				

	UFR Uala III	or remivable as
	per the	requirements"
Remarks	Intervals of	10 sec" Not
	second by se	econd. No data
	sync with other	er parameters.

III. The online price bid of the eligible tenderer mentioned above will be opened on 26.10.2024 @ 11.00 am at the e-tender portal.

Dr. Shibulal AGeneral Manager