



## KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)
Thycaud.P.O., Thiruvananthapuram-14.
Tele Fax No. 0471-2945647
0471-2945646, 0471-2945600

E-mail id: <a href="mailto:eptenders.kmscl@kerala.gov.in">eptenders.kmscl@kerala.gov.in</a>
CIN: U24233KL2007SGC021616
GSTIN: 32AADCK4029M1ZK

www.kmscl.kerala.gov.in

Dated: 10.01.2024

**EP Division** 

No: KMSCL/EP/17267/2023

## NOTICE

Sub:- Tender No.T512/ 2023 - Results of demonstration/ technical evaluation & Date of Price Bid Opening – Anesthesia Work Station - reg.

Ref:- 1. Tender Inviting Notice published dated 04.10.2023

- 2. Technical bid opening held on 30.10.2023 @ 11.00 A M
- 3. Eligible list published on 03.11.2023
- 4. Demonstration/ document evaluation of Anesthesia Work Station conducted on 16.12.2023 & 06.01.2024
- I. The reports of technical document evaluation/ demonstration of Anesthesia Work Station as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

SI. No	Equipment	Name of the Selecte Bidder	Selected Models
1.	Anesthesia Work Station	M/s Sree Gokulam	WATO EX-35/Shenzhen
		Healthcare Pvt. Ltd.	Mindray

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

## A. M/s Universal Agencies

Rejected model/ OEM: URJA/ REUBENS HOSPITECH

Reason for rejection:

Clause no	Technical specifications	Deviations
6	Should have monitoring facility of airway pressure, tidal volume, frequency, oxygen concentration and AGM with modular integration.	ETCO2 not integrated
11	Should supply pin index yokes for connecting cylinders for O2-1No, N2O -1No and 1no for Air through pipeline.	Connection for Air through pipeline not available
	Should have safety certificate from a competent authority CE issued by a notified body registered in the European	
25	commission / FDA (US)/ STQC CE Certificate/ STQC S Certificate or valid detailed electrical and functional safety tes report from ERTL. Copy of the certificate test report shall be produced along with the technical bid	No valid certificate
38	Should have Medical grade IP44 or above rated power cord to match D-Type plug The protective earthling resistance (PER	

III. The online price bid of the eligible tenderer mentioned above will be opened on 6.01.2024 @ 11.00 am at the e-tender portal.

**Dr. Shibulal A**General Manager