

**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/11988/2023

Dated: 23.05.2024

NOTICE**Sub:-** Tender No.T499/ 2023 - Cancellation Notice - various equipment - reg.**Ref:-** 1. Tender Inviting Notice published dated 05.07.2023

2. Technical bid opening on 24.07.2023 & 07.08.2023@ 11.00 A M

3. Reports of demonstration/ document evaluation of Electrosurgical Unit with vessel sealing system for General Surgery, Urology and Gynecology received on 16.05.2024

4. Reports of demonstration/ document evaluation of Whole Body Phototherapy Unit received on 22.05.2024

I. The report of technical document evaluation / demonstration of as per reference no 3rd & 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers participated in the technical evaluation are as follows;

Sl .No	Equipment Name	Name of the selected bidder	Selected model
1.	Whole Body Phototherapy Unit	M/s Kerala Surgical Equipment Co.	SPIEGEL SERIES /DERMAINDIA

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below:

A. KMSCL/EP/T499/433/2023: Whole Body Phototherapy Unit**1. M/s Universal Agencies**

Rejected model: Square 12+12 ANB/Healkeren Medtronic's

Reason for rejection:

Clause no	Technical Specification	Deviations noted
5	Should have On site calibration.	Not meeting the specification
11	Should have Cooling system (Fan).	Not meeting the specification

2. M/s Lakshmy Agency

Rejected model: SPIEGEL SERIES/DERMAINDIA

Reason for rejection:

Clause no	Technical Specification	Deviations noted
	Comments	Manufactures authorization not submitted.