



**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [eptenders.kmscl@kerala.gov.in](mailto:eptenders.kmscl@kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

**EP Division**

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

**No: KMSCL/EP/1720/2024**

**Dated: 26.06.2024**

**NOTICE**

**Sub:-** Tender No.T527/ 2024 - Cancellation Notice – Automated Coagulation Analyzer - reg.

- Ref:-**
1. Tender Inviting Notice published dated 01.02.2024
  2. Technical bid opening held on 17.02.2024 @ 11.00 A M
  3. Reports of demonstration/ document evaluation of Automated Coagulation Analyzer received on 05.03.2024

I. The reports of technical document evaluation/ demonstration of Automated Coagulation Analyzer as per reference no. 3rd cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

SI No	Equipment Name	Selected Bidders	Model/OEM
1.	Automated Coagulation Analyzer	M/s LabX	STA Satellite Max/ DIAGNOSTICA

II. The other tenderers participated in the tender are rejected due to the reason mentioned against their name below;

A. M/s Transasia Bio Medicals Ltd

Rejected Model: ECL-760/ Erba Lachema

Reason for rejection;

Clause No	Technical specification	Compliance of the model
5	Tests parameters required-PT, APTT, TT, ATIII, Fibrinogen, Lupus Anticoagulant Screen and Confirm, Anti-Xa (UFH & LMWH), D-dimer, Chromogenic Protein C, Chromogenic Free Protein S. Factor VIII and Factor IX.	Anti-Xa (UFH & LMWH) not mentioned in catalogue
6	Reagents and Consumables should be manufactured by same manufacturer as that of the Instrument to ensure standardised protocol for any tests.	Above missing parameter - no clarification
7	Clot Based Assays are to be insensitive to high bilirubin and Lipemic Samples.	Endogenous interference mentioned
15	Instrument should be able to accommodate minimum of 200 cuvettes on board.	Only 72 cuvettes

II. Due to administrative reason the tender for KMSCL/EP/T527/1276/2024: Automated Coagulation Analyzer is cancelled.

**Dr. Shibulal A**  
General Manager