



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)
Thycaud.P.O., Thiruvananthapuram-14.
Tele Fax No. 0471-2945647
0471-2945646, 0471-2945600

E-mail id: eptenders.kmscl@kerala.gov.in CIN: U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

www.kmscl.kerala.gov.in

EP Division

No: KMSCL/EP/662/2024

Dated: 14.06.2024

NOTICE

Sub:- Tender No.T522/ 2024 - Results of demonstration/ technical evaluation & Date of Price Bid Opening - X Ray film Lobby LED Three Film - reg.

- Ref:- 1. Tender Inviting Notice published dated 04.10.2023
 - 2. Technical bid opening held on 27.10.2023 @ 11.00 A M
 - 3. Eligible list published on 28.11.2024
 - 4. Demonstration/ document evaluation of X Ray film Lobby LED Three Film conducted on 01.06.2024
- I. The reports of technical document evaluation/ demonstration of X Ray film Lobby LED Three Film as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

SI No	Equipment	Name of the Selected Bidder	Selected Models
1.	X Ray film Lobby LED Three Film	TECHNOLOGIES	3LEDXRAYVIEW/ C CUBE
		M/s. UNIVERSAL AGENCIES	SPARX XVB14/ HARDIK MEDI-TECH

- II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;
- M/s. Alan Medical & Laboratory Products
 Rejected model: Lyfkeeran/A P Enterprises

 Reason for rejection;

Clause No	Technical Specification	Deviations Noted
8	LED Lamps should provide a uniform level of illumination across the entire front panel diffuser and should be controlled by electronic step-less dimming controls to provide flicker free dimming from maximum brightness to off.	Not an uniform level of illumination

2. M/s. LAKSHMI AGENCY

Rejected model; Gulati Industries

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
8	electronic step-less dimming controls toil	Not an uniform level of Ilumination
	brightness to off.	

3. M/s. x ray medico surgical engineering ancillary

Rejected model: XMSEA/ X-ray Medico

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
	Comments	Not a single unit, separate power supply

4. M/s. Trivandrum Surgicals and Lab Supplies

Rejected model. DUKLS/BHARAT SURGICAL COMPANY

Reason for rejection;

Clause No	Technical Specification	Deviations Noted
	Comments	Not a single unit,seperate power supply

III. The online price bid of the eligible tenderer mentioned above will be opened on 18.06.2024 @ 11.00 am at the e-tender portal.

Dr. Shibulal A
General Manager